

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062922\
 Data File : VY009382.D
 Acq On : 29 Jun 2022 19:00
 Operator : KP/MD
 Sample : N3488-02
 Misc : 4.55g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-9-8.5-10.5-20220623

Quant Time: Jun 30 05:13:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 17:35:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	4705	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	6430	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	6476	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	2369	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	2326	61.722	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 123.440%	
35) Dibromofluoromethane	7.716	113	647	18.282	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 36.560%#	
50) Toluene-d8	10.179	98	6470	59.630	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 119.260%	
62) 4-Bromofluorobenzene	12.477	95	1450	28.177	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 56.360%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062922\
Data File : VY009382.D
Acq On : 29 Jun 2022 19:00
Operator : KP/MD
Sample : N3488-02
Misc : 4.55g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-9-8.5-10.5-20220623

Quant Time: Jun 30 05:13:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 17:35:34 2022
Response via : Initial Calibration

