

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062922\
 Data File : VY009376.D
 Acq On : 29 Jun 2022 16:31
 Operator : KP/MD
 Sample : N3462-03
 Misc : 4.37g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-8-9-11-20220622

Quant Time: Jun 29 17:45:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 17:35:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	146709	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	252115	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	240975	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	107387	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	67910	57.792	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.580%
35) Dibromofluoromethane	7.709	113	58827	42.395	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	84.800%
50) Toluene-d8	10.172	98	228945	53.815	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.620%
62) 4-Bromofluorobenzene	12.477	95	72822	36.092	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	72.180%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062922\
Data File : VY009376.D
Acq On : 29 Jun 2022 16:31
Operator : KP/MD
Sample : N3462-03
Misc : 4.37g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-8-9-11-20220622

Quant Time: Jun 29 17:45:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 17:35:34 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY009376.D
Acq: 29 Jun 2022 16:31




Instrument :
MSVOA_Y
ClientSampleId :
SB-8-9-11-20220622


Tgt Ion:168 Resp: 146709
Ion Ratio Lower Upper
168 100
99 50.2 46.9 70.3



#33
1,2-Dichloroethane-d4
Concen: 57.792 ug/l
RT: 8.136 min Scan# 1036
Delta R.T. 0.000 min
Lab File: VY009376.D
Acq: 29 Jun 2022 16:31

Tgt Ion: 65 Resp: 67910
Ion Ratio Lower Upper
65 100
67 51.8 0.0 99.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1126

Delta R.T. -0.000 min

Lab File: VY009376.D

Acq: 29 Jun 2022 16:31

Instrument :

MSVOA_Y

ClientSampleId :

SB-8-9-11-20220622





Tgt Ion:114 Resp: 252115

Ion Ratio Lower Upper

114 100

63 19.8 0.0 43.2

88 14.9 0.0 31.8



#35

Dibromofluoromethane

Concen: 42.395 ug/l

RT: 7.709 min Scan# 966

Delta R.T. -0.001 min

Lab File: VY009376.D

Acq: 29 Jun 2022 16:31




Tgt Ion:113 Resp: 58827


Ion Ratio Lower Upper

113 100

111 103.1 82.6 124.0

192 20.8 17.6 26.4









#50
Toluene-d8
Concen: 53.815 ug/l
RT: 10.172 min Scan# 11
Delta R.T. -0.007 min
Lab File: VY009376.D
Acq: 29 Jun 2022 16:31




Instrument :
MSVOA_Y
ClientSampleId :
SB-8-9-11-20220622


Tgt Ion: 98 Resp: 228945
Ion Ratio Lower Upper
98 100
100 65.0 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 36.092 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. -0.000 min
Lab File: VY009376.D
Acq: 29 Jun 2022 16:31

Tgt Ion: 95 Resp: 72822
Ion Ratio Lower Upper
95 100
174 87.8 0.0 177.0
176 83.8 0.0 175.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.489 min Scan# 11

Delta R.T. -0.001 min

Lab File: VY009376.D

Acq: 29 Jun 2022 16:31

Instrument :

MSVOA_Y

ClientSampleId :

SB-8-9-11-20220622





Tgt Ion:117 Resp: 240975

Ion Ratio Lower Upper

117 100

82 52.8 43.0 64.4

119 33.5 25.4 38.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.422 min Scan# 1903

Delta R.T. -0.000 min

Lab File: VY009376.D

Acq: 29 Jun 2022 16:31

Tgt Ion:152 Resp: 107387

Ion Ratio Lower Upper

152 100

115 56.5 28.2 84.7

150 161.3 0.0 347.6

