

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY063021\
 Data File : VY005294.D
 Acq On : 30 Jun 2021 13:33
 Operator : SY/MD
 Sample : VY0630SBL01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0630SBL01

Quant Time: Jul 01 06:36:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062221S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 07:57:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	261082	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	386568	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	352610	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	173505	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	126791	58.521	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.040%
35) Dibromofluoromethane	7.722	113	117102	51.365	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.720%
50) Toluene-d8	10.185	98	465431	51.050	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.100%
62) 4-Bromofluorobenzene	12.483	95	148173	48.772	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.540%
Target Compounds						
16) Acetone	3.960	43	8238	16.459	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY063021\
Data File : VY005294.D
Acq On : 30 Jun 2021 13:33
Operator : SY/MD
Sample : VY0630SBL01
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0630SBL01

Quant Time: Jul 01 06:36:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062221S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 07:57:52 2021
Response via : Initial Calibration

