

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY063021\
 Data File : VY005302.D
 Acq On : 30 Jun 2021 16:51
 Operator : SY/MD
 Sample : M2902-01
 Misc : 5.05G/5ML/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CL-01-063021

Quant Time: Jul 01 06:39:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062221S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 07:57:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	223409	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	364740	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	368262	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	190643	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	126763	68.374	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 136.740%	
35) Dibromofluoromethane	7.722	113	115464	53.677	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 107.360%	
50) Toluene-d8	10.185	98	452525	52.605	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 105.200%	
62) 4-Bromofluorobenzene	12.483	95	161717	56.415	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 112.840%	
Target Compounds						
16) Acetone	3.960	43	13795	32.209	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY063021\
Data File : VY005302.D
Acq On : 30 Jun 2021 16:51
Operator : SY/MD
Sample : M2902-01
Misc : 5.05G/5ML/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CL-01-063021

Quant Time: Jul 01 06:39:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062221S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 07:57:52 2021
Response via : Initial Calibration

