

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY063022\
 Data File : VY009398.D
 Acq On : 30 Jun 2022 17:10
 Operator : KP/MD
 Sample : N3488-09
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-12-3-5-20220623

Quant Time: Jul 01 03:48:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 17:35:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	138441	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	238231	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	225235	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	100467	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	62130	56.031	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.060%
35) Dibromofluoromethane	7.710	113	56010	42.718	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	85.440%
50) Toluene-d8	10.173	98	209215	52.043	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.080%
62) 4-Bromofluorobenzene	12.477	95	66836	35.055	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	70.120%
Target Compounds						
16) Acetone	3.942	43	5639	15.485	ug/l	Qvalue # 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY063022\
Data File : VY009398.D
Acq On : 30 Jun 2022 17:10
Operator : KP/MD
Sample : N3488-09
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-12-3-5-20220623

Quant Time: Jul 01 03:48:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 17:35:34 2022
Response via : Initial Calibration

