

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY063022\
 Data File : VY009400.D
 Acq On : 30 Jun 2022 17:56
 Operator : KP/MD
 Sample : N3488-11
 Misc : 4.77g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-7-0-2-20220623

Quant Time: Jul 01 03:48:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 17:35:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	138357	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	228212	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	219082	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	100252	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	61573	55.562	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.120%
35) Dibromofluoromethane	7.709	113	52555	41.842	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	83.680%
50) Toluene-d8	10.172	98	201589	52.348	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.700%
62) 4-Bromofluorobenzene	12.477	95	63864	34.967	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	69.940%
Target Compounds						
16) Acetone	3.942	43	5433	14.928	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY063022\
Data File : VY009400.D
Acq On : 30 Jun 2022 17:56
Operator : KP/MD
Sample : N3488-11
Misc : 4.77g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-7-0-2-20220623

Quant Time: Jul 01 03:48:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 17:35:34 2022
Response via : Initial Calibration

