

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY063022\
 Data File : VY009407.D
 Acq On : 30 Jun 2022 20:40
 Operator : KP/MD
 Sample : N3550-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-02-06292022

Quant Time: Jul 01 03:49:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 17:35:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	128539	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	215808	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	210203	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	94126	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	57739	56.082	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.160%
35) Dibromofluoromethane	7.716	113	51516	43.373	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	86.740%
50) Toluene-d8	10.179	98	193064	53.015	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.040%
62) 4-Bromofluorobenzene	12.477	95	62780	36.349	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	72.700%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY063022\
Data File : VY009407.D
Acq On : 30 Jun 2022 20:40
Operator : KP/MD
Sample : N3550-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HR-02-06292022

Quant Time: Jul 01 03:49:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 17:35:34 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY009407.D
Acq: 30 Jun 2022 20:40




Instrument :
MSVOA_Y
ClientSampleId :
HR-02-06292022


Tgt Ion:168 Resp: 128539
Ion Ratio Lower Upper
168 100
99 51.1 46.9 70.3



#33
1,2-Dichloroethane-d4
Concen: 56.082 ug/l
RT: 8.136 min Scan# 1036
Delta R.T. 0.000 min
Lab File: VY009407.D
Acq: 30 Jun 2022 20:40

Tgt Ion: 65 Resp: 57739
Ion Ratio Lower Upper
65 100
67 51.3 0.0 99.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1126

Delta R.T. -0.000 min

Lab File: VY009407.D

Acq: 30 Jun 2022 20:40

Instrument :

MSVOA_Y

ClientSampleId :

HR-02-06292022





Tgt Ion:114 Resp: 215808

Ion Ratio Lower Upper

114 100

63 19.1 0.0 43.2

88 14.1 0.0 31.8



#35

Dibromofluoromethane

Concen: 43.373 ug/l

RT: 7.716 min Scan# 967

Delta R.T. 0.006 min

Lab File: VY009407.D

Acq: 30 Jun 2022 20:40




Tgt Ion:113 Resp: 51516


Ion Ratio Lower Upper

113 100

111 102.5 82.6 124.0

192 20.2 17.6 26.4









#50
Toluene-d8
Concen: 53.015 ug/l
RT: 10.179 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY009407.D
Acq: 30 Jun 2022 20:40




Instrument :
MSVOA_Y
ClientSampleId :
HR-02-06292022


Tgt Ion: 98 Resp: 193064
Ion Ratio Lower Upper
98 100
100 64.6 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 36.349 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. -0.000 min
Lab File: VY009407.D
Acq: 30 Jun 2022 20:40

Tgt Ion: 95 Resp: 62780
Ion Ratio Lower Upper
95 100
174 88.6 0.0 177.0
176 86.9 0.0 175.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.489 min Scan# 11

Delta R.T. -0.001 min

Lab File: VY009407.D

Acq: 30 Jun 2022 20:40

Instrument :

MSVOA_Y

ClientSampleId :

HR-02-06292022





Tgt Ion:117 Resp: 210203

Ion Ratio Lower Upper

117 100

82 52.1 43.0 64.4

119 31.9 25.4 38.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.422 min Scan# 1903

Delta R.T. -0.000 min

Lab File: VY009407.D

Acq: 30 Jun 2022 20:40

Tgt Ion:152 Resp: 94126

Ion Ratio Lower Upper

152 100

115 56.2 28.2 84.7

150 158.9 0.0 347.6

