



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY063022\
Data File : VY009389.D
Acq On : 30 Jun 2022 10:16
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
Title : SW846 8260
Last Update : Wed Jun 29 17:35:34 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.0	9772	PASS
75	95	30	60	52.6	24515	PASS
95	95	100	100	100.0	46571	PASS
96	95	5	9	7.3	3393	PASS
173	174	0.00	2	1.5	629	PASS
174	95	50	100	91.1	42411	PASS
175	174	5	9	7.3	3096	PASS
176	174	95	101	97.7	41429	PASS
177	176	5	9	7.2	2977	PASS