

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070121\
 Data File : VY005319.D
 Acq On : 01 Jul 2021 17:12
 Operator : SY/MD
 Sample : M2904-04
 Misc : 5.41G/5ML/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-3

Quant Time: Jul 02 05:17:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062221S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 07:57:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	226011	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	366607	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	377100	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	198558	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	126482	67.437	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 134.880%	
35) Dibromofluoromethane	7.728	113	114433	52.927	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 105.860%	
50) Toluene-d8	10.185	98	454715	52.590	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 105.180%	
62) 4-Bromofluorobenzene	12.483	95	165009	57.271	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 114.540%	
Target Compounds						
						Qvalue
16) Acetone	3.954	43	4998	11.535	ug/l	99
20) Methylene Chloride	4.722	84	28600	4.469	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070121\
Data File : VY005319.D
Acq On : 01 Jul 2021 17:12
Operator : SY/MD
Sample : M2904-04
Misc : 5.41G/5ML/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-3

Quant Time: Jul 02 05:17:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062221S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 07:57:52 2021
Response via : Initial Calibration

