



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070121\
Data File : VY005307.D
Acq On : 01 Jul 2021 09:50
Operator : SY/MD
Sample : BFB
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062221S.M
Title : SW846 8260
Last Update : Wed Jun 23 07:57:52 2021



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.1	22952	PASS
75	95	30	60	50.8	61197	PASS
95	95	100	100	100.0	120448	PASS
96	95	5	9	6.7	8032	PASS
173	174	0.00	2	0.8	915	PASS
174	95	50	100	96.2	115904	PASS
175	174	5	9	7.6	8803	PASS
176	174	95	101	97.7	113235	PASS
177	176	5	9	7.8	8785	PASS