

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070122\
 Data File : VY009420.D
 Acq On : 01 Jul 2022 14:20
 Operator : KP/MD
 Sample : N3548-01
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 V-1

Quant Time: Jul 05 02:45:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 17:35:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	84477	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	143066	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	138405	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	63122	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	40629	60.046	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	120.100%
35) Dibromofluoromethane	7.710	113	34983	44.428	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	88.860%
50) Toluene-d8	10.173	98	125575	52.016	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.040%
62) 4-Bromofluorobenzene	12.477	95	40286	35.185	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	70.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070122\
Data File : VY009420.D
Acq On : 01 Jul 2022 14:20
Operator : KP/MD
Sample : N3548-01
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
V-1

Quant Time: Jul 05 02:45:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 17:35:34 2022
Response via : Initial Calibration

