

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070122\
 Data File : VY009421.D
 Acq On : 01 Jul 2022 14:43
 Operator : KP/MD
 Sample : N3557-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-02-063022

Quant Time: Jul 05 02:45:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 17:35:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	127384	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	217685	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	202580	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	96261	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	57274	56.135	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.260%
35) Dibromofluoromethane	7.710	113	50701	42.318	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	84.640%
50) Toluene-d8	10.173	98	193755	52.746	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.500%
62) 4-Bromofluorobenzene	12.477	95	61491	35.296	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	70.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070122\
Data File : VY009421.D
Acq On : 01 Jul 2022 14:43
Operator : KP/MD
Sample : N3557-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-02-063022

Quant Time: Jul 05 02:45:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 17:35:34 2022
Response via : Initial Calibration

