

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070122\
 Data File : VY009421.D
 Acq On : 01 Jul 2022 14:43
 Operator : KP/MD
 Sample : N3557-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-02-063022

Quant Time: Jul 05 02:45:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 17:35:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	127384	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	217685	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	202580	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	96261	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	57274	56.135	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.260%
35) Dibromofluoromethane	7.710	113	50701	42.318	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	84.640%
50) Toluene-d8	10.173	98	193755	52.746	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.500%
62) 4-Bromofluorobenzene	12.477	95	61491	35.296	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	70.600%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070122\
Data File : VY009421.D
Acq On : 01 Jul 2022 14:43
Operator : KP/MD
Sample : N3557-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-02-063022

Quant Time: Jul 05 02:45:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 17:35:34 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY009421.D
Acq: 01 Jul 2022 14:43




Instrument :
MSVOA_Y
ClientSampleId :
OR-02-063022


Tgt Ion:168 Resp: 127384
Ion Ratio Lower Upper
168 100
99 51.8 46.9 70.3



#33
1,2-Dichloroethane-d4
Concen: 56.135 ug/l
RT: 8.137 min Scan# 1036
Delta R.T. 0.001 min
Lab File: VY009421.D
Acq: 01 Jul 2022 14:43

Tgt Ion: 65 Resp: 57274
Ion Ratio Lower Upper
65 100
67 51.2 0.0 99.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1126

Delta R.T. 0.000 min

Lab File: VY009421.D

Acq: 01 Jul 2022 14:43

Instrument :

MSVOA_Y

ClientSampleId :

OR-02-063022





Tgt Ion:114 Resp: 217685

Ion Ratio Lower Upper

114 100

63 19.6 0.0 43.2

88 14.4 0.0 31.8



#35

Dibromofluoromethane

Concen: 42.318 ug/l

RT: 7.710 min Scan# 966

Delta R.T. -0.000 min

Lab File: VY009421.D

Acq: 01 Jul 2022 14:43




Tgt Ion:113 Resp: 50701


Ion Ratio Lower Upper

113 100

111 105.2 82.6 124.0

192 21.0 17.6 26.4









#50
Toluene-d8
Concen: 52.746 ug/l
RT: 10.173 min Scan# 11
Delta R.T. -0.006 min
Lab File: VY009421.D
Acq: 01 Jul 2022 14:43




Instrument :
MSVOA_Y
ClientSampleId :
OR-02-063022


Tgt Ion: 98 Resp: 193755
Ion Ratio Lower Upper
98 100
100 65.2 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 35.296 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. 0.000 min
Lab File: VY009421.D
Acq: 01 Jul 2022 14:43

Tgt Ion: 95 Resp: 61491
Ion Ratio Lower Upper
95 100
174 92.2 0.0 177.0
176 88.3 0.0 175.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.490 min Scan# 11

Delta R.T. -0.000 min

Lab File: VY009421.D

Acq: 01 Jul 2022 14:43

Instrument :

MSVOA_Y

ClientSampleId :

OR-02-063022





Tgt Ion:117 Resp: 202580

Ion Ratio Lower Upper

117 100

82 51.9 43.0 64.4

119 32.2 25.4 38.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.422 min Scan# 1903

Delta R.T. 0.000 min

Lab File: VY009421.D

Acq: 01 Jul 2022 14:43

Tgt Ion:152 Resp: 96261

Ion Ratio Lower Upper

152 100

115 54.0 28.2 84.7

150 155.8 0.0 347.6

