

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY070220\
 Data File : VY003084.D
 Acq On : 02 Jul 2020 14:47
 Operator : SY/MD
 Sample : L3176-03
 Misc : 2.32G/5ML/MSVOA Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 STONE-2

Quant Time: Jul 03 03:20:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y062920S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 30 04:11:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.80	168	290644	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.69	114	474275	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.49	117	457104	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.42	152	232035	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.15	65	164997	57.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	114.32%	
35) Dibromofluoromethane	7.73	113	159286	53.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.68%	
50) Toluene-d8	10.18	98	572074	51.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.80%	
62) 4-Bromofluorobenzene	12.48	95	201456	51.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.84%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY070220\
Data File : VY003084.D
Acq On : 02 Jul 2020 14:47
Operator : SY/MD
Sample : L3176-03
Misc : 2.32G/5ML/MSVOA Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
STONE-2

Quant Time: Jul 03 03:20:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\82Y062920S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 30 04:11:35 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.80 min Scan# 994

Delta R.T. -0.00 min

Lab File: VY003084.D

Acq: 02 Jul 2020 14:47

Instrument :

MSVOA_Y

ClientSampleId :


STONE-2

Tot Ion:168 Resp: 290644

Ion Ratio Lower Upper

168 100

99 50.2 39.2 58.8



Abundance Ion 167.90 (167.60 to 168.60): VY003002.D (-983) (-)

Ion 98.90 (98.60 to 99.60): VY003084.D (-945) (-)

7.80

150000



100000

50000

0

7.70 7.80 7.90

Time-->



#33

1,2-Dichloroethane-d4

Concen: 57.158 ug/l

RT: 8.15 min Scan# 1052

Delta R.T. 0.01 min

Lab File: VY003084.D


Acq: 02 Jul 2020 14:47

Tgt Ion: 65 Resp: 164997

Ion Ratio Lower Upper

65 100

67 52.7 0.0 105.6



Abundance Ion 64.90 (64.60 to 65.60): VY003002.D (-1040) (-)

Ion 66.90 (66.60 to 67.60): VY003084.D (-1002) (-)

8.15

80000

60000


40000


20000

0

8.10 8.20

Time-->









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.69 min Scan# 1141
 Delta R.T. -0.00 min
 Lab File: VY003084.D
 Acq: 02 Jul 2020 14:47




Instrument :
 MSVOA_Y
 ClientSampleId :
 STONE-2


Tot Ion:114 Resp: 474275
 Ion Ratio Lower Upper
 114 100
 63 20.4 0.0 40.6
 88 15.4 0.0 30.0



#35
 Dibromofluoromethane
 Concen: 53.838 ug/l
 RT: 7.73 min Scan# 983
 Delta R.T. 0.01 min
 Lab File: VY003084.D
 Acq: 02 Jul 2020 14:47

Tgt Ion:113 Resp: 159286
 Ion Ratio Lower Upper
 113 100
 111 102.8 82.6 123.8
 192 22.5 18.1 27.1





#50

Toluene-d8

Concen: 51.902 ug/l

RT: 10.18 min Scan# 1385

Delta R.T. -0.00 min

Lab File: VY003084.D

Acq: 02 Jul 2020 14:47

Instrument :

MSVOA_Y

ClientSampleId :





STONE-2

Tot Ion: 98 Resp: 572074

Ion Ratio Lower Upper

98 100

100 63.5 51.1 76.7



#62

4-Bromofluorobenzene

Concen: 51.924 ug/l

RT: 12.48 min Scan# 1762

Delta R.T. -0.00 min

Lab File: VY003084.D

Acq: 02 Jul 2020 14:47




Tgt Ion: 95 Resp: 201456


Ion Ratio Lower Upper

95 100

174 95.5 0.0 184.8

176 91.8 0.0 183.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.49 min Scan# 1600
Delta R.T. -0.00 min
Lab File: VY003084.D
Acq: 02 Jul 2020 14:47

Instrument :
MSVOA_Y
ClientSampleId :
STONE-2

Tot Ion:117 Resp: 457104
Ion Ratio Lower Upper
117 100
82 52.7 43.0 64.6
119 32.0 25.8 38.8




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VY0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.42 min Scan# 1917
Delta R.T. -0.00 min
Lab File: VY003084.D
Acq: 02 Jul 2020 14:47

Tgt Ion:152 Resp: 232035
Ion Ratio Lower Upper
152 100
115 54.8 27.3 81.9
150 156.4 0.0 343.4



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

