

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070221\
 Data File : VY005343.D
 Acq On : 02 Jul 2021 15:32
 Operator : SY/MD
 Sample : M2933-03
 Misc : 5.03G/5ML/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 3

Quant Time: Jul 05 01:05:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062221S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 07:57:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	136	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	293	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	165	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.43
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	78	69.112	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 138.220%	
35) Dibromofluoromethane	0.000	113	0	0.000	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	= 0.000%#	
50) Toluene-d8	10.179	98	373	53.977	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 107.960%	
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	25 - 144	Recovery	= 0.000%#	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070221\
Data File : VY005343.D
Acq On : 02 Jul 2021 15:32
Operator : SY/MD
Sample : M2933-03
Misc : 5.03G/5ML/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
3

Quant Time: Jul 05 01:05:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062221S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 07:57:52 2021
Response via : Initial Calibration

