

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070323\
 Data File : VY014544.D
 Acq On : 03 Jul 2023 18:43
 Operator : KP/MD
 Sample : 03475-03
 Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ALDENE-COMP-2

Quant Time: Jul 04 03:04:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y063023S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 30 19:02:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	283821	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	449572	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	435329	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	211804	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	156044	55.268	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.540%
35) Dibromofluoromethane	7.710	113	145608	52.595	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.180%
50) Toluene-d8	10.179	98	531774	48.590	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.180%
62) 4-Bromofluorobenzene	12.477	95	176530	46.452	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	92.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070323\
Data File : VY014544.D
Acq On : 03 Jul 2023 18:43
Operator : KP/MD
Sample : 03475-03
Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ALDENE-COMP-2

Quant Time: Jul 04 03:04:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y063023S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 30 19:02:11 2023
Response via : Initial Calibration

