

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070323\
 Data File : VY014546.D
 Acq On : 03 Jul 2023 19:31
 Operator : KP/MD
 Sample : 03488-03
 Misc : 8.55g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-1-GRAB

Quant Time: Jul 04 03:05:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y063023S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 30 19:02:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	267011	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	426794	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	419777	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	201877	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	160289	60.346	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	120.700%
35) Dibromofluoromethane	7.715	113	143494	54.597	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.200%
50) Toluene-d8	10.178	98	512680	49.345	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.700%
62) 4-Bromofluorobenzene	12.477	95	168052	46.581	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	93.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070323\
Data File : VY014546.D
Acq On : 03 Jul 2023 19:31
Operator : KP/MD
Sample : 03488-03
Misc : 8.55g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-1-GRAB

Quant Time: Jul 04 03:05:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y063023S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 30 19:02:11 2023
Response via : Initial Calibration

