

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070323\
 Data File : VY014547.D
 Acq On : 03 Jul 2023 19:54
 Operator : KP/MD
 Sample : 03507-01
 Misc : 5.14g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CP-1

Quant Time: Jul 04 03:06:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y063023S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 30 19:02:11 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	280730	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	444709	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	444219	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	222377	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	152534	54.620	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 109.240%		
35) Dibromofluoromethane	7.716	113	143486	52.395	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 104.780%		
50) Toluene-d8	10.179	98	537953	49.692	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	= 99.380%		
62) 4-Bromofluorobenzene	12.477	95	189555	50.425	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	= 100.840%		
Target Compounds						
						Qvalue
20) Methylene Chloride	4.704	84	39692	8.285	ug/l #	82
29) Tetrahydrofuran	7.350	42	2452	2.944	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070323\
Data File : VY014547.D
Acq On : 03 Jul 2023 19:54
Operator : KP/MD
Sample : 03507-01
Misc : 5.14g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CP-1

Quant Time: Jul 04 03:06:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y063023S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 30 19:02:11 2023
Response via : Initial Calibration

