

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070323\
 Data File : VY014548.D
 Acq On : 03 Jul 2023 20:18
 Operator : KP/MD
 Sample : 03507-04
 Misc : 5.13g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CP-2

Quant Time: Jul 04 03:06:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y063023S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 30 19:02:11 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.782	168	294226	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	461584	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	461099	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	230971	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	161292	55.106	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	110.220%	
35) Dibromofluoromethane	7.709	113	148535	52.256	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	104.520%	
50) Toluene-d8	10.178	98	553834	49.288	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	98.580%	
62) 4-Bromofluorobenzene	12.477	95	200628	51.419	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	102.840%	
Target Compounds						
						Qvalue
16) Acetone	3.942	43	47583	52.057	ug/l #	83
29) Tetrahydrofuran	7.350	42	3215	3.683	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070323\
Data File : VY014548.D
Acq On : 03 Jul 2023 20:18
Operator : KP/MD
Sample : 03507-04
Misc : 5.13g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CP-2

Quant Time: Jul 04 03:06:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y063023S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 30 19:02:11 2023
Response via : Initial Calibration

