

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070323\
 Data File : VY014550.D
 Acq On : 03 Jul 2023 21:05
 Operator : KP/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Jul 04 03:07:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y063023S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 30 19:02:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	310038	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	481197	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	460018	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	222153	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	159481	51.709	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 103.420%	
35) Dibromofluoromethane	7.716	113	151670	51.184	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.360%	
50) Toluene-d8	10.179	98	569319	48.601	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 97.200%	
62) 4-Bromofluorobenzene	12.477	95	189350	46.551	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 93.100%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.704	84	44359	8.453	ug/l	# 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070323\
Data File : VY014550.D
Acq On : 03 Jul 2023 21:05
Operator : KP/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Jul 04 03:07:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y063023S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 30 19:02:11 2023
Response via : Initial Calibration

