

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070522\
 Data File : VY009436.D
 Acq On : 05 Jul 2022 14:44
 Operator : KP/MD
 Sample : N3568-01
 Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-05-07012022

Quant Time: Jul 06 04:50:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 17:35:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	1971	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	3184	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	3198	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	1086	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	1129	71.515	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	143.020%
35) Dibromofluoromethane	7.704	113	294	16.777	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	33.560%#
50) Toluene-d8	10.179	98	2918	54.310	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	108.620%
62) 4-Bromofluorobenzene	12.483	95	909	35.673	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	71.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070522\
Data File : VY009436.D
Acq On : 05 Jul 2022 14:44
Operator : KP/MD
Sample : N3568-01
Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-05-07012022

Quant Time: Jul 06 04:50:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 17:35:34 2022
Response via : Initial Calibration

