



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070522\
 Data File : VY009427.D
 Acq On : 05 Jul 2022 08:55
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
 Title : SW846 8260
 Last Update : Wed Jun 29 17:35:34 2022



AutoFind: Scans 1747, 1748, 1749; Background Corrected with Scan 1739

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.0	12971	PASS
75	95	30	60	53.8	31672	PASS
95	95	100	100	100.0	58891	PASS
96	95	5	9	7.0	4138	PASS
173	174	0.00	2	1.2	615	PASS
174	95	50	100	87.8	51731	PASS
175	174	5	9	7.5	3862	PASS
176	174	95	101	96.8	50075	PASS
177	176	5	9	6.5	3268	PASS