

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070523\
 Data File : VY014557.D
 Acq On : 05 Jul 2023 14:45
 Operator : KP/MD
 Sample : 03498-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 NB-07-063023

Quant Time: Jul 05 15:01:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y063023S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 30 19:02:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	234996	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	376533	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	365530	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	170349	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	148208	63.399	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	126.800%
35) Dibromofluoromethane	7.710	113	127126	54.826	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.660%
50) Toluene-d8	10.173	98	454702	49.607	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.220%
62) 4-Bromofluorobenzene	12.477	95	149143	46.858	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	93.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070523\
Data File : VY014557.D
Acq On : 05 Jul 2023 14:45
Operator : KP/MD
Sample : 03498-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
NB-07-063023

Quant Time: Jul 05 15:01:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y063023S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 30 19:02:11 2023
Response via : Initial Calibration

