

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070523\
 Data File : VY014560.D
 Acq On : 05 Jul 2023 15:56
 Operator : KP/MD
 Sample : 03475-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ALDENE-COMP-1

Quant Time: Jul 06 01:33:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y063023S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 30 19:02:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	224739	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	366301	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	354461	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	171774	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	141324	63.213	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	126.420%
35) Dibromofluoromethane	7.710	113	123468	54.736	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.480%
50) Toluene-d8	10.173	98	444600	49.859	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.720%
62) 4-Bromofluorobenzene	12.477	95	147046	47.490	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	94.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070523\
Data File : VY014560.D
Acq On : 05 Jul 2023 15:56
Operator : KP/MD
Sample : 03475-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ALDENE-COMP-1

Quant Time: Jul 06 01:33:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y063023S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 30 19:02:11 2023
Response via : Initial Calibration

