

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070523\
 Data File : VY014572.D
 Acq On : 05 Jul 2023 20:41
 Operator : KP/MD
 Sample : 03514-01
 Misc : 2.65g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CP-4

Quant Time: Jul 06 01:38:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y063023S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 30 19:02:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	222352	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	343354	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	352838	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	174216	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	118905	53.756	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.520%
35) Dibromofluoromethane	7.710	113	110477	52.250	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.500%
50) Toluene-d8	10.179	98	392058	46.906	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	93.820%
62) 4-Bromofluorobenzene	12.477	95	157374	54.222	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	108.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070523\
Data File : VY014572.D
Acq On : 05 Jul 2023 20:41
Operator : KP/MD
Sample : 03514-01
Misc : 2.65g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CP-4

Quant Time: Jul 06 01:38:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y063023S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 30 19:02:11 2023
Response via : Initial Calibration

