

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070622\
 Data File : VY009451.D
 Acq On : 06 Jul 2022 14:08
 Operator : KP/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Jul 07 01:49:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 17:35:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	245034	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	403939	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	380546	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	184854	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	106825	54.430	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 108.860%	
35) Dibromofluoromethane	7.710	113	93326	41.979	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 83.960%	
50) Toluene-d8	10.173	98	357863	52.501	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	= 105.000%	
62) 4-Bromofluorobenzene	12.477	95	133938	41.432	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	= 82.860%	
Target Compounds						
					Qvalue	
39) Methylcyclohexane	9.179	83	5876	1.277	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070622\
Data File : VY009451.D
Acq On : 06 Jul 2022 14:08
Operator : KP/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Jul 07 01:49:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 17:35:34 2022
Response via : Initial Calibration

