

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070622\
 Data File : VY009452.D
 Acq On : 06 Jul 2022 14:31
 Operator : KP/MD
 Sample : N3579-01
 Misc : 4.85g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OK-03-070522

Quant Time: Jul 07 01:49:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 17:35:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	103716	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	167482	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	154025	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	69666	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	43697	52.601	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	105.200%
35) Dibromofluoromethane	7.710	113	40068	43.468	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	86.940%
50) Toluene-d8	10.173	98	149424	52.871	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	105.740%
62) 4-Bromofluorobenzene	12.477	95	50684	37.813	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	75.620%
Target Compounds						
39) Methylcyclohexane	9.173	83	3201	1.678	ug/l	Qvalue # 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070622\
Data File : VY009452.D
Acq On : 06 Jul 2022 14:31
Operator : KP/MD
Sample : N3579-01
Misc : 4.85g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OK-03-070522

Quant Time: Jul 07 01:49:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 17:35:34 2022
Response via : Initial Calibration

