

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070622\
 Data File : VY009453.D
 Acq On : 06 Jul 2022 14:54
 Operator : KP/MD
 Sample : N3581-01
 Misc : 5.15g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 AU-6-070522

Quant Time: Jul 07 01:50:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 17:35:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	224347	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	381541	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	354069	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	173402	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	99428	55.332	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.660%
35) Dibromofluoromethane	7.710	113	91646	43.643	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	87.280%
50) Toluene-d8	10.173	98	333112	51.739	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	103.480%
62) 4-Bromofluorobenzene	12.477	95	120250	39.381	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	78.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070622\
Data File : VY009453.D
Acq On : 06 Jul 2022 14:54
Operator : KP/MD
Sample : N3581-01
Misc : 5.15g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
AU-6-070522

Quant Time: Jul 07 01:50:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 17:35:34 2022
Response via : Initial Calibration

