

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070622\
 Data File : VY009456.D
 Acq On : 06 Jul 2022 16:05
 Operator : KP/MD
 Sample : N3584-05
 Misc : 4.24g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RB-21199

Quant Time: Jul 07 01:50:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 17:35:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	192550	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	327221	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	295338	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	122466	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	83043	53.845	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 107.700%	
35) Dibromofluoromethane	7.710	113	76498	42.477	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 84.960%	
50) Toluene-d8	10.173	98	288543	52.256	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	= 104.520%	
62) 4-Bromofluorobenzene	12.477	95	94086	35.927	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	= 71.860%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	4.942	59	34178	108.933	ug/l #	92
39) Methylcyclohexane	9.173	83	4060	1.089	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070622\
Data File : VY009456.D
Acq On : 06 Jul 2022 16:05
Operator : KP/MD
Sample : N3584-05
Misc : 4.24g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RB-21199

Quant Time: Jul 07 01:50:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 17:35:34 2022
Response via : Initial Calibration

