

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070622\
 Data File : VY009445.D
 Acq On : 06 Jul 2022 11:17
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 07 01:47:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 17:35:34 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.412	0.348	15.5	99	0.00
3 P	Chloromethane	0.448	0.369	17.6	96	0.00
4 C	Vinyl Chloride	0.518	0.457	11.8#	97	0.00
5 T	Bromomethane	0.332	0.280	15.7	104	0.00
6 T	Chloroethane	0.276	0.249	9.8	101	0.00
7 T	Trichlorofluoromethane	0.687	0.637	7.3	102	0.00
8 T	Diethyl Ether	0.269	0.259	3.7	111	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.493	0.448	9.1	105	0.00
10 T	Methyl Iodide	0.541	0.538	0.6	97	0.00
11 T	Tert butyl alcohol	0.091	0.059	35.2#	104	0.00
12 CM	1,1-Dichloroethene	0.449	0.406	9.6#	104	0.00
13 T	Acrolein	0.033	0.025	24.2	94	0.00
14 T	Allyl chloride	0.824	0.772	6.3	102	0.00
15 T	Acrylonitrile	0.152	0.143	5.9	102	0.00
16 T	Acetone	0.132	0.117	11.4	106	0.00
17 T	Carbon Disulfide	1.250	1.001	19.9	91	0.00
18 T	Methyl Acetate	0.397	0.344	13.4	110	0.00
19 T	Methyl tert-butyl Ether	1.348	1.350	-0.1	107	-0.01
20 T	Methylene Chloride	0.814	0.526	35.4#	99	0.00
21 T	trans-1,2-Dichloroethene	0.538	0.493	8.4	98	0.00
22 T	Diisopropyl ether	1.738	1.868	-7.5	106	0.00
23 T	Vinyl Acetate	1.089	1.170	-7.4	106	0.00
24 P	1,1-Dichloroethane	0.977	0.955	2.3	104	0.00
25 T	2-Butanone	0.209	0.199	4.8	104	0.00
26 T	2,2-Dichloropropane	0.925	0.873	5.6	105	0.00
27 T	cis-1,2-Dichloroethene	0.625	0.612	2.1	103	0.00
28 T	Bromochloromethane	0.420	0.443	-5.5	99	0.00
29 T	Tetrahydrofuran	0.134	0.133	0.7	103	0.00
30 C	Chloroform	1.004	1.006	-0.2#	106	0.00
31 T	Cyclohexane	0.901	0.791	12.2	98	0.00
32 T	1,1,1-Trichloroethane	0.899	0.874	2.8	105	0.00
33 S	1,2-Dichloroethane-d4	0.400	0.393	1.8	106	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	104	0.00
35 S	Dibromofluoromethane	0.275	0.269	2.2	104	0.00
36 T	1,1-Dichloropropene	0.481	0.455	5.4	101	0.00
37 T	Ethyl Acetate	0.285	0.279	2.1	104	0.00
38 T	Carbon Tetrachloride	0.503	0.491	2.4	103	0.00
39 T	Methylcyclohexane	0.569	0.533	6.3	98	0.00
40 TM	Benzene	1.407	1.366	2.9	101	0.00
41 T	Methacrylonitrile	0.161	0.155	3.7	84	0.00
42 TM	1,2-Dichloroethane	0.404	0.397	1.7	104	0.00
43 T	Isopropyl Acetate	0.524	0.538	-2.7	107	0.00
44 TM	Trichloroethene	0.391	0.378	3.3	105	0.00
45 C	1,2-Dichloropropane	0.356	0.364	-2.2#	104	0.00
46 T	Dibromomethane	0.204	0.201	1.5	103	0.00
47 T	Bromodichloromethane	0.484	0.482	0.4	105	0.00
48 T	Methyl methacrylate	0.226	0.235	-4.0	108	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070622\
 Data File : VY009445.D
 Acq On : 06 Jul 2022 11:17
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 07 01:47:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 17:35:34 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	102	0.00
50 S	Toluene-d8	0.844	0.830	1.7	100	0.00
51 T	4-Methyl-2-Pentanone	0.270	0.283	-4.8	108	0.00
52 CM	Toluene	0.888	0.901	-1.5#	103	0.00
53 T	t-1,3-Dichloropropene	0.484	0.529	-9.3	106	0.00
54 T	cis-1,3-Dichloropropene	0.578	0.579	-0.2	102	0.00
55 T	1,1,2-Trichloroethane	0.272	0.298	-9.6	109	0.00
56 T	Ethyl methacrylate	0.347	0.410	-18.2	109	0.00
57 T	1,3-Dichloropropane	0.481	0.492	-2.3	103	0.00
58 T	2-Chloroethyl Vinyl ether	0.192	0.213	-10.9	109	0.00
59 T	2-Hexanone	0.187	0.196	-4.8	108	0.00
60 T	Dibromochloromethane	0.344	0.360	-4.7	105	0.00
61 T	1,2-Dibromoethane	0.273	0.274	-0.4	105	0.00
62 S	4-Bromofluorobenzene	0.400	0.428	-7.0	109	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	103	0.00
64 T	Tetrachloroethene	0.465	0.462	0.6	109	0.00
65 PM	Chlorobenzene	1.070	1.058	1.1	104	0.00
66 T	1,1,1,2-Tetrachloroethane	0.392	0.402	-2.6	109	0.00
67 C	Ethyl Benzene	1.856	1.887	-1.7#	105	0.00
68 T	m/p-Xylenes	0.742	0.736	0.8	106	0.00
69 T	o-Xylene	0.700	0.722	-3.1	108	0.00
70 T	Styrene	1.160	1.212	-4.5	108	0.00
71 P	Bromoform	0.247	0.255	-3.2	110	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	0.00
73 T	Isopropylbenzene	3.693	3.563	3.5	107	0.00
74 T	N-amyl acetate	0.969	0.967	0.2	109	0.00
75 P	1,1,2,2-Tetrachloroethane	0.715	0.678	5.2	109	0.00
76 T	1,2,3-Trichloropropane	0.586	0.498	15.0	104	0.00
77 T	Bromobenzene	0.880	0.845	4.0	108	0.00
78 T	n-propylbenzene	4.492	4.380	2.5	111	0.00
79 T	2-Chlorotoluene	2.536	2.462	2.9	110	0.00
80 T	1,3,5-Trimethylbenzene	3.088	3.016	2.3	109	0.00
81 T	trans-1,4-Dichloro-2-butene	0.263	0.254	3.4	107	0.00
82 T	4-Chlorotoluene	2.635	2.531	3.9	109	0.00
83 T	tert-Butylbenzene	2.595	2.680	-3.3	110	0.00
84 T	1,2,4-Trimethylbenzene	2.935	2.991	-1.9	109	0.00
85 T	sec-Butylbenzene	3.908	4.012	-2.7	111	0.00
86 T	p-Isopropyltoluene	3.286	3.319	-1.0	110	0.00
87 T	1,3-Dichlorobenzene	1.739	1.693	2.6	111	0.00
88 T	1,4-Dichlorobenzene	1.790	1.690	5.6	110	0.00
89 T	n-Butylbenzene	3.088	3.120	-1.0	124	0.00
90 T	Hexachloroethane	0.652	0.650	0.3	115	0.00
91 T	1,2-Dichlorobenzene	1.577	1.534	2.7	114	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.120	0.113	5.8	112	0.00
93 T	1,2,4-Trichlorobenzene	0.936	0.950	-1.5	119	0.00
94 T	Hexachlorobutadiene	0.526	0.516	1.9	114	0.00
95 T	Naphthalene	1.875	1.967	-4.9	117	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070622\
Data File : VY009445.D
Acq On : 06 Jul 2022 11:17
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jul 07 01:47:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 17:35:34 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.814	0.842	-3.4	117	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6