

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070623\
 Data File : VY014580.D
 Acq On : 06 Jul 2023 11:55
 Operator : KP/MD
 Sample : 03515-01
 Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 LEONIA-COMP

Quant Time: Jul 07 06:11:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y063023S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 30 19:02:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	233414	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	372358	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	350791	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	167650	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	138191	59.515	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.020%
35) Dibromofluoromethane	7.710	113	124110	54.125	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.260%
50) Toluene-d8	10.179	98	426015	46.998	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.000%
62) 4-Bromofluorobenzene	12.477	95	140061	44.498	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	89.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070623\
Data File : VY014580.D
Acq On : 06 Jul 2023 11:55
Operator : KP/MD
Sample : 03515-01
Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
LEONIA-COMP

Quant Time: Jul 07 06:11:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y063023S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 30 19:02:11 2023
Response via : Initial Calibration

