

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070623\
 Data File : VY014587.D
 Acq On : 06 Jul 2023 14:42
 Operator : KP/MD
 Sample : 03522-02
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VOC

Quant Time: Jul 07 06:13:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y063023S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 30 19:02:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	210150	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	336408	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	332173	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	153972	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	131921	63.104	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	126.200%
35) Dibromofluoromethane	7.710	113	115812	55.904	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.800%
50) Toluene-d8	10.179	98	410190	50.088	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.180%
62) 4-Bromofluorobenzene	12.477	95	131998	46.418	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	92.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070623\
Data File : VY014587.D
Acq On : 06 Jul 2023 14:42
Operator : KP/MD
Sample : 03522-02
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VOC

Quant Time: Jul 07 06:13:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y063023S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 30 19:02:11 2023
Response via : Initial Calibration

