

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070723\
 Data File : VY014603.D
 Acq On : 07 Jul 2023 17:43
 Operator : KP/MD
 Sample : 03555-07
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ-203

Quant Time: Jul 08 03:47:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y063023S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 30 19:02:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	215816	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	351116	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	338252	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	146951	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	137790	64.181	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	128.360%
35) Dibromofluoromethane	7.710	113	119101	55.083	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.160%
50) Toluene-d8	10.179	98	425997	49.839	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.680%
62) 4-Bromofluorobenzene	12.477	95	133074	44.836	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	89.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070723\
Data File : VY014603.D
Acq On : 07 Jul 2023 17:43
Operator : KP/MD
Sample : 03555-07
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-203

Quant Time: Jul 08 03:47:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y063023S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 30 19:02:11 2023
Response via : Initial Calibration

