



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070723\
Data File : VY014594.D
Acq On : 07 Jul 2023 08:46
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y063023S.M
Title : SW846 8260
Last Update : Fri Jun 30 19:02:11 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1739

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	17.1	12553	PASS
75	95	30	60	48.5	35501	PASS
95	95	100	100	100.0	73224	PASS
96	95	5	9	6.9	5047	PASS
173	174	0.00	2	1.4	999	PASS
174	95	50	100	100.0	73200	PASS
175	174	5	9	7.5	5481	PASS
176	174	95	101	95.6	69989	PASS
177	176	5	9	7.0	4875	PASS