

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv070820\
 Data File : VY003134.D
 Acq On : 08 Jul 2020 13:16
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 09 02:41:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y062920S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 30 04:11:35 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	90	0.00
2 T	Dichlorodifluoromethane	0.379	0.366	3.4	87	0.00
3 P	Chloromethane	0.555	0.497	10.5	89	0.00
4 C	Vinyl Chloride	0.541	0.544	-0.6#	91	0.00
5 T	Bromomethane	0.367	0.381	-3.8	98	0.00
6 T	Chloroethane	0.344	0.358	-4.1	95	0.00
7 T	Trichlorofluoromethane	0.763	0.797	-4.5	94	0.00
8 T	Diethyl Ether	0.262	0.262	0.0	90	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.456	0.495	-8.6	98	0.00
10 T	Methyl Iodide	0.565	0.619	-9.6	94	0.00
11 T	Tert butyl alcohol	0.051	0.047	7.8	84	0.00
12 CM	1,1-Dichloroethene	0.434	0.458	-5.5#	96	0.00
13 T	Acrolein	0.045	0.042	6.7	90	0.00
14 T	Allyl chloride	0.757	0.684	9.6	82	0.00
15 T	Acrylonitrile	0.135	0.131	3.0	85	0.00
16 T	Acetone	0.117	0.102	12.8	74	0.00
17 T	Carbon Disulfide	1.386	1.338	3.5	87	0.00
18 T	Methyl Acetate	0.316	0.289	8.5	83	0.00
19 T	Methyl tert-butyl Ether	1.264	1.293	-2.3	90	0.00
20 T	Methylene Chloride	0.654	0.518	20.8	89	0.00
21 T	trans-1,2-Dichloroethene	0.491	0.516	-5.1	97	0.00
22 T	Diisopropyl ether	1.623	1.557	4.1	87	0.00
23 T	Vinyl Acetate	1.085	1.022	5.8	83	0.00
24 P	1,1-Dichloroethane	0.887	0.882	0.6	91	0.00
25 T	2-Butanone	0.186	0.168	9.7	79	0.00
26 T	2,2-Dichloropropane	0.806	0.822	-2.0	94	0.00
27 T	cis-1,2-Dichloroethene	0.550	0.578	-5.1	96	0.00
28 T	Bromochloromethane	0.394	0.400	-1.5	86	0.00
29 T	Tetrahydrofuran	0.124	0.113	8.9	80	0.00
30 C	Chloroform	0.878	0.922	-5.0#	95	0.00
31 T	Cyclohexane	0.873	0.813	6.9	88	0.00
32 T	1,1,1-Trichloroethane	0.802	0.833	-3.9	95	0.00
33 S	1,2-Dichloroethane-d4	0.497	0.464	6.6	81	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	88	0.00
35 S	Dibromofluoromethane	0.312	0.329	-5.4	88	0.00
36 T	1,1-Dichloropropene	0.482	0.507	-5.2	93	0.00
37 T	Ethyl Acetate	0.290	0.279	3.8	82	0.00
38 T	Carbon Tetrachloride	0.501	0.543	-8.4	95	0.00
39 T	Methylcyclohexane	0.597	0.638	-6.9	93	0.00
40 TM	Benzene	1.319	1.399	-6.1	93	0.00
41 T	Methacrylonitrile	0.155	0.178	-14.8	86	0.00
42 TM	1,2-Dichloroethane	0.417	0.428	-2.6	89	0.00
43 T	Isopropyl Acetate	0.551	0.530	3.8	81	0.00
44 TM	Trichloroethene	0.383	0.437	-14.1	101	0.00
45 C	1,2-Dichloropropane	0.355	0.364	-2.5#	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv070820\
 Data File : VY003134.D
 Acq On : 08 Jul 2020 13:16
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 09 02:41:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y062920S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 30 04:11:35 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.198	0.212	-7.1	94	0.00
47 T	Bromodichloromethane	0.485	0.517	-6.6	93	0.00
48 T	Methyl methacrylate	0.251	0.240	4.4	81	0.00
49 T	1,4-Dioxane	0.002	0.003	-50.0#	97	0.00
50 S	Toluene-d8	1.162	1.195	-2.8	86	0.00
51 T	4-Methyl-2-Pentanone	0.285	0.283	0.7	84	0.00
52 CM	Toluene	0.830	0.899	-8.3#	95	0.00
53 T	t-1,3-Dichloropropene	0.509	0.535	-5.1	91	0.00
54 T	cis-1,3-Dichloropropene	0.582	0.614	-5.5	92	0.00
55 T	1,1,2-Trichloroethane	0.275	0.298	-8.4	93	0.00
56 T	Ethyl methacrylate	0.399	0.427	-7.0	90	0.00
57 T	1,3-Dichloropropane	0.483	0.512	-6.0	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.204	0.212	-3.9	83	0.00
59 T	2-Hexanone	0.199	0.195	2.0	82	0.00
60 T	Dibromochloromethane	0.352	0.389	-10.5	96	0.00
61 T	1,2-Dibromoethane	0.271	0.296	-9.2	95	0.00
62 S	4-Bromofluorobenzene	0.409	0.421	-2.9	86	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.423	0.519	-22.7	106	0.00
65 PM	Chlorobenzene	0.986	1.099	-11.5	98	0.00
66 T	1,1,1,2-Tetrachloroethane	0.385	0.432	-12.2	97	0.00
67 C	Ethyl Benzene	1.804	1.997	-10.7#	97	0.00
68 T	m/p-Xylenes	0.671	0.753	-12.2	97	0.00
69 T	o-Xylene	0.633	0.703	-11.1	96	0.00
70 T	Styrene	1.081	1.223	-13.1	97	0.00
71 P	Bromoform	0.257	0.289	-12.5	94	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
73 T	Isopropylbenzene	3.272	3.664	-12.0	98	0.00
74 T	N-amyl acetate	1.065	1.043	2.1	83	0.00
75 P	1,1,2,2-Tetrachloroethane	0.660	0.681	-3.2	90	0.00
76 T	1,2,3-Trichloropropane	0.495	0.534	-7.9	93	0.00
77 T	Bromobenzene	0.839	0.918	-9.4	97	0.00
78 T	n-propylbenzene	3.935	4.353	-10.6	97	0.00
79 T	2-Chlorotoluene	2.197	2.392	-8.9	96	0.00
80 T	1,3,5-Trimethylbenzene	2.746	3.058	-11.4	98	0.00
81 T	trans-1,4-Dichloro-2-butene	0.254	0.264	-3.9	88	0.00
82 T	4-Chlorotoluene	2.287	2.512	-9.8	96	0.00
83 T	tert-Butylbenzene	2.411	2.711	-12.4	98	0.00
84 T	1,2,4-Trimethylbenzene	2.767	3.088	-11.6	98	0.00
85 T	sec-Butylbenzene	3.323	3.771	-13.5	99	0.00
86 T	p-Isopropyltoluene	3.079	3.502	-13.7	99	0.00
87 T	1,3-Dichlorobenzene	1.566	1.757	-12.2	98	0.00
88 T	1,4-Dichlorobenzene	1.545	1.735	-12.3	98	0.00
89 T	n-Butylbenzene	2.946	3.348	-13.6	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv070820\
Data File : VY003134.D
Acq On : 08 Jul 2020 13:16
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jul 09 02:41:24 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y062920S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 30 04:11:35 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.611	0.660	-8.0	94	0.00
91 T	1,2-Dichlorobenzene	1.405	1.568	-11.6	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.117	0.113	3.4	81	0.00
93 T	1,2,4-Trichlorobenzene	1.044	1.129	-8.1	94	0.00
94 T	Hexachlorobutadiene	0.659	0.762	-15.6	101	0.00
95 T	Naphthalene	1.907	1.964	-3.0	87	0.00
96 T	1,2,3-Trichlorobenzene	0.907	0.933	-2.9	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6