

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY070920\
 Data File : VY003145.D
 Acq On : 09 Jul 2020 15:46
 Operator : SY/MD
 Sample : L3255-01
 Misc : 5.84G/5ML/MSVOA Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CHRT-57

Quant Time: Jul 10 10:00:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y062920S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 30 04:11:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.80	168	299322	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.69	114	474143	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.49	117	443706	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.43	152	202227	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.14	65	161171	54.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.42%	
35) Dibromofluoromethane	7.72	113	155882	52.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.40%	
50) Toluene-d8	10.18	98	563844	51.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.34%	
62) 4-Bromofluorobenzene	12.48	95	188549	48.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.22%	
Target Compounds						
16) Acetone	3.96	43	15300	21.821	ug/l	91
25) 2-Butanone	7.00	43	7018	6.301	ug/l #	77
92) 1,2-Dibromo-3-Chloropropan	14.35	75	523	1.106	ug/l	73
93) 1,2,4-Trichlorobenzene	15.01	180	11720	2.775	ug/l	96
94) Hexachlorobutadiene	15.12	225	4169	1.565	ug/l	92
95) Naphthalene	15.23	128	25773	3.341	ug/l	99
96) 1,2,3-Trichlorobenzene	15.42	180	12677	3.456	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY070920\
Data File : VY003145.D
Acq On : 09 Jul 2020 15:46
Operator : SY/MD
Sample : L3255-01
Misc : 5.84G/5ML/MSVOA Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CHRT-57

Quant Time: Jul 10 10:00:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\82Y062920S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 30 04:11:35 2020
Response via : Initial Calibration

