



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070921\
 Data File : VY005390.D
 Acq On : 09 Jul 2021 08:57
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070921S.M
 Title : SW846 8260
 Last Update : Sat Jul 10 08:11:17 2021



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.3	17501	PASS
75	95	30	60	48.8	52212	PASS
95	95	100	100	100.0	107043	PASS
96	95	5	9	6.9	7337	PASS
173	174	0.00	2	0.4	406	PASS
174	95	50	100	95.7	102429	PASS
175	174	5	9	8.1	8300	PASS
176	174	95	101	96.4	98736	PASS
177	176	5	9	7.0	6939	PASS