

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071023\
 Data File : VY014614.D
 Acq On : 10 Jul 2023 13:42
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Jul 11 03:18:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y063023S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 30 19:02:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	243229	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	377537	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	370525	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	185574	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	132246	54.656	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.320%	
35) Dibromofluoromethane	7.716	113	125261	53.878	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 107.760%	
50) Toluene-d8	10.179	98	460607	50.117	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 100.240%	
62) 4-Bromofluorobenzene	12.477	95	157854	49.463	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 98.920%	
Target Compounds						
20) Methylene Chloride	4.710	84	47107	13.485	ug/l	Qvalue # 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071023\
Data File : VY014614.D
Acq On : 10 Jul 2023 13:42
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Jul 11 03:18:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y063023S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 30 19:02:11 2023
Response via : Initial Calibration

