

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071322\
 Data File : VY009555.D
 Acq On : 13 Jul 2022 13:43
 Operator : KP/MD
 Sample : N3597-05
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 20220510-COMPOSITE-E

Quant Time: Jul 14 02:44:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 08 11:00:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	135050	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	201140	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	175675	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	79568	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	40529	41.502	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	83.000%
35) Dibromofluoromethane	7.710	113	47871	45.232	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.460%
50) Toluene-d8	10.173	98	120694	39.964	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	79.920%
62) 4-Bromofluorobenzene	12.477	95	59632	39.916	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	79.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071322\
Data File : VY009555.D
Acq On : 13 Jul 2022 13:43
Operator : KP/MD
Sample : N3597-05
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
20220510-COMPOSITE-E

Quant Time: Jul 14 02:44:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070722S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 08 11:00:56 2022
Response via : Initial Calibration

