

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071322\
 Data File : VY009556.D
 Acq On : 13 Jul 2022 14:06
 Operator : KP/MD
 Sample : N3597-06RE
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 20220511-FIELD-DUPLICATE-COMPRE

Quant Time: Jul 14 02:44:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 08 11:00:56 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 07/14/2022
 Supervised By : Mahesh Dadoda 07/14/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	130283	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	202761	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	165658	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	68671	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	42935	46.305	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	92.600%		
35) Dibromofluoromethane	7.710	113	45374	42.530	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	85.060%		
50) Toluene-d8	10.173	98	113217	36.515	ug/l	0.00
Spiked Amount 50.000	Range 78 - 125		Recovery =	73.040%#		
62) 4-Bromofluorobenzene	12.477	95	52224	34.678	ug/l	0.00
Spiked Amount 50.000	Range 50 - 146		Recovery =	69.360%		
Target Compounds					Qvalue	
16) Acetone	3.948	43	3572	11.271	ug/l	98
25) 2-Butanone	6.966	43	3152m	6.148	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071322\
Data File : VY009556.D
Acq On : 13 Jul 2022 14:06
Operator : KP/MD
Sample : N3597-06RE
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
20220511-FIELD-DUPLICATE-COMP

Quant Time: Jul 14 02:44:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070722S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 08 11:00:56 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 07/14/2022
Supervised By : Mahesh Dadoda 07/14/2022

