

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071322\
 Data File : VY009560.D
 Acq On : 13 Jul 2022 17:11
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 14 02:45:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 08 11:00:56 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	79	0.00
2 T	Dichlorodifluoromethane	0.383	1.086	-183.6#	224#	0.00
3 P	Chloromethane	0.428	0.441	-3.0	77	0.00
4 C	Vinyl Chloride	0.488	0.543	-11.3#	81	0.00
5 T	Bromomethane	0.350	0.366	-4.6	84	-0.01
6 T	Chloroethane	0.285	0.313	-9.8	87	0.00
7 T	Trichlorofluoromethane	0.651	0.713	-9.5	84	0.00
8 T	Diethyl Ether	0.249	0.248	0.4	76	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.441	0.427	3.2	75	0.00
10 T	Methyl Iodide	0.523	0.556	-6.3	75	0.00
11 T	Tert butyl alcohol	0.077	0.053	31.2#	69	0.00
12 CM	1,1-Dichloroethene	0.409	0.412	-0.7#	75	0.00
13 T	Acrolein	0.031	0.027	12.9	54	0.00
14 T	Allyl chloride	0.704	0.707	-0.4	71	0.00
15 T	Acrylonitrile	0.137	0.140	-2.2	76	0.00
16 T	Acetone	0.122	0.113	7.4	55	0.00
17 T	Carbon Disulfide	1.151	1.171	-1.7	66	0.00
18 T	Methyl Acetate	0.357	0.329	7.8	78	0.00
19 T	Methyl tert-butyl Ether	1.238	1.287	-4.0	80	0.00
20 T	Methylene Chloride	0.685	0.543	20.7	75	0.00
21 T	trans-1,2-Dichloroethene	0.479	0.485	-1.3	74	0.00
22 T	Diisopropyl ether	1.496	1.708	-14.2	80	0.00
23 T	Vinyl Acetate	0.957	1.134	-18.5	79	0.00
24 P	1,1-Dichloroethane	0.844	0.857	-1.5	77	0.00
25 T	2-Butanone	0.197	0.203	-3.0	70	0.00
26 T	2,2-Dichloropropane	0.832	0.828	0.5	78	0.00
27 T	cis-1,2-Dichloroethene	0.564	0.572	-1.4	78	0.00
28 T	Bromochloromethane	0.355	0.382	-7.6	80	0.00
29 T	Tetrahydrofuran	0.127	0.138	-8.7	79	0.00
30 C	Chloroform	0.923	0.916	0.8#	79	0.00
31 T	Cyclohexane	0.820	0.794	3.2	71	0.00
32 T	1,1,1-Trichloroethane	0.814	0.825	-1.4	79	0.00
33 S	1,2-Dichloroethane-d4	0.427	0.344	19.4	82	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	77	0.00
35 S	Dibromofluoromethane	0.263	0.240	8.7	81	0.00
36 T	1,1-Dichloropropene	0.428	0.453	-5.8	76	0.00
37 T	Ethyl Acetate	0.268	0.299	-11.6	81	0.00
38 T	Carbon Tetrachloride	0.458	0.482	-5.2	77	0.00
39 T	Methylcyclohexane	0.514	0.542	-5.4	71	0.00
40 TM	Benzene	1.258	1.308	-4.0	75	0.00
41 T	Methacrylonitrile	0.139	0.170	-22.3	85	-0.01
42 TM	1,2-Dichloroethane	0.378	0.404	-6.9	80	0.00
43 T	Isopropyl Acetate	0.501	0.555	-10.8	81	0.00
44 TM	Trichloroethene	0.355	0.373	-5.1	77	0.00
45 C	1,2-Dichloropropane	0.318	0.335	-5.3#	76	0.00
46 T	Dibromomethane	0.187	0.200	-7.0	79	0.00
47 T	Bromodichloromethane	0.445	0.468	-5.2	80	0.00
48 T	Methyl methacrylate	0.216	0.240	-11.1	78	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071322\
 Data File : VY009560.D
 Acq On : 13 Jul 2022 17:11
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 14 02:45:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 08 11:00:56 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	80	0.00
50 S	Toluene-d8	0.904	0.713	21.1	81	0.00
51 T	4-Methyl-2-Pentanone	0.263	0.306	-16.3	83	0.00
52 CM	Toluene	0.801	0.884	-10.4#	77	0.00
53 T	t-1,3-Dichloropropene	0.469	0.518	-10.4	79	0.00
54 T	cis-1,3-Dichloropropene	0.522	0.559	-7.1	77	0.00
55 T	1,1,2-Trichloroethane	0.270	0.295	-9.3	80	0.00
56 T	Ethyl methacrylate	0.358	0.417	-16.5	79	0.00
57 T	1,3-Dichloropropane	0.444	0.483	-8.8	79	0.00
58 T	2-Chloroethyl Vinyl ether	0.175	0.188	-7.4	74	0.00
59 T	2-Hexanone	0.183	0.220	-20.2	80	0.00
60 T	Dibromochloromethane	0.326	0.362	-11.0	81	0.00
61 T	1,2-Dibromoethane	0.258	0.281	-8.9	79	0.00
62 S	4-Bromofluorobenzene	0.371	0.405	-9.2	87	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	79	0.00
64 T	Tetrachloroethene	0.418	0.390	6.7	70	0.00
65 PM	Chlorobenzene	0.982	1.040	-5.9	80	0.00
66 T	1,1,1,2-Tetrachloroethane	0.369	0.394	-6.8	82	0.00
67 C	Ethyl Benzene	1.714	1.842	-7.5#	80	0.00
68 T	m/p-Xylenes	0.669	0.744	-11.2	82	0.00
69 T	o-Xylene	0.641	0.722	-12.6	82	0.00
70 T	Styrene	1.070	1.246	-16.4	85	0.00
71 P	Bromoform	0.235	0.270	-14.9	85	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
73 T	Isopropylbenzene	3.419	3.466	-1.4	83	0.00
74 T	N-amyl acetate	0.973	1.086	-11.6	94	0.00
75 P	1,1,2,2-Tetrachloroethane	0.685	0.718	-4.8	87	0.00
76 T	1,2,3-Trichloropropane	0.529	0.504	4.7	74	0.00
77 T	Bromobenzene	0.832	0.836	-0.5	84	0.00
78 T	n-propylbenzene	4.131	4.266	-3.3	84	0.00
79 T	2-Chlorotoluene	2.313	2.400	-3.8	85	0.00
80 T	1,3,5-Trimethylbenzene	2.858	2.930	-2.5	84	0.00
81 T	trans-1,4-Dichloro-2-butene	0.259	0.270	-4.2	85	0.00
82 T	4-Chlorotoluene	2.424	2.527	-4.2	87	0.00
83 T	tert-Butylbenzene	2.543	2.641	-3.9	85	0.00
84 T	1,2,4-Trimethylbenzene	2.845	2.978	-4.7	86	0.00
85 T	sec-Butylbenzene	3.745	3.949	-5.4	87	0.00
86 T	p-Isopropyltoluene	3.122	3.315	-6.2	88	0.00
87 T	1,3-Dichlorobenzene	1.643	1.704	-3.7	88	0.00
88 T	1,4-Dichlorobenzene	1.663	1.709	-2.8	88	0.00
89 T	n-Butylbenzene	2.915	3.093	-6.1	88	0.00
90 T	Hexachloroethane	0.613	0.627	-2.3	86	0.00
91 T	1,2-Dichlorobenzene	1.506	1.583	-5.1	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.120	0.123	-2.5	87	0.00
93 T	1,2,4-Trichlorobenzene	0.918	0.922	-0.4	86	0.00
94 T	Hexachlorobutadiene	0.505	0.493	2.4	85	0.00
95 T	Naphthalene	1.899	1.937	-2.0	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071322\
Data File : VY009560.D
Acq On : 13 Jul 2022 17:11
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jul 14 02:45:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070722S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 08 11:00:56 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.795	0.799	-0.5	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6