

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071322\
 Data File : VY009549.D
 Acq On : 13 Jul 2022 10:22
 Operator : KP/MD
 Sample : VY0713SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0713SBL01

Quant Time: Jul 14 02:42:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 08 11:00:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	146052	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	235423	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	211819	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	99493	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	51555	50.128	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	100.260%
35) Dibromofluoromethane	7.709	113	56361	45.499	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.000%
50) Toluene-d8	10.179	98	153366	44.218	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	88.440%
62) 4-Bromofluorobenzene	12.477	95	76512	43.757	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	87.520%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071322\
Data File : VY009549.D
Acq On : 13 Jul 2022 10:22
Operator : KP/MD
Sample : VY0713SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0713SBL01

Quant Time: Jul 14 02:42:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070722S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 08 11:00:56 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY009549.D
Acq: 13 Jul 2022 10:22




Instrument :
MSVOA_Y
ClientSampleId :
VY0713SBL01


Tgt Ion:168 Resp: 146052
Ion Ratio Lower Upper
168 100
99 52.8 46.9 70.3



#33
1,2-Dichloroethane-d4
Concen: 50.128 ug/l
RT: 8.136 min Scan# 1036
Delta R.T. 0.000 min
Lab File: VY009549.D
Acq: 13 Jul 2022 10:22

Tgt Ion: 65 Resp: 51555
Ion Ratio Lower Upper
65 100
67 52.5 0.0 99.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.685 min Scan# 1125
Delta R.T. 0.006 min
Lab File: VY009549.D
Acq: 13 Jul 2022 10:22




Instrument :
MSVOA_Y
ClientSampleId :
VY0713SBL01


Tgt Ion:114 Resp: 235423
Ion Ratio Lower Upper
114 100
63 19.3 0.0 43.2
88 14.4 0.0 31.8



#35
Dibromofluoromethane
Concen: 45.499 ug/l
RT: 7.709 min Scan# 966
Delta R.T. -0.001 min
Lab File: VY009549.D
Acq: 13 Jul 2022 10:22

Tgt Ion:113 Resp: 56361
Ion Ratio Lower Upper
113 100
111 102.1 82.6 124.0
192 21.2 17.6 26.4









#50
Toluene-d8
Concen: 44.218 ug/l
RT: 10.179 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY009549.D
Acq: 13 Jul 2022 10:22




Instrument :
MSVOA_Y
ClientSampleId :
VY0713SBL01


Tgt Ion: 98 Resp: 153366
Ion Ratio Lower Upper
98 100
100 65.9 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 43.757 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. -0.000 min
Lab File: VY009549.D
Acq: 13 Jul 2022 10:22

Tgt Ion: 95 Resp: 76512
Ion Ratio Lower Upper
95 100
174 92.3 0.0 177.0
176 87.1 0.0 175.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY009549.D
Acq: 13 Jul 2022 10:22

Instrument :
MSVOA_Y
ClientSampleId :
VY0713SBL01

Tgt Ion:117 Resp: 211819
Ion Ratio Lower Upper
117 100
82 52.1 43.0 64.4
119 33.1 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.422 min Scan# 1903
Delta R.T. -0.000 min
Lab File: VY009549.D
Acq: 13 Jul 2022 10:22

Tgt Ion:152 Resp: 99493
Ion Ratio Lower Upper
152 100
115 58.1 28.2 84.7
150 158.6 0.0 347.6

