

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071322\
 Data File : VY009560.D
 Acq On : 13 Jul 2022 17:11
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By : John Carlone 07/14/2022
 Supervised By : Mahesh Dadoda 07/14/2022

Quant Time: Jul 14 02:45:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 08 11:00:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	153684	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.679	114	239217	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	225256	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	123960	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	52888	48.683	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	97.360%		
35) Dibromofluoromethane	7.710	113	57523	45.701	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	91.400%		
50) Toluene-d8	10.173	98	170576	49.317	ug/l	0.00
Spiked Amount 50.000	Range 78 - 125		Recovery =	98.640%		
62) 4-Bromofluorobenzene	12.477	95	96846	54.508	ug/l	0.00
Spiked Amount 50.000	Range 50 - 146		Recovery =	109.020%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.900	85	166886	141.603	ug/l	100
3) Chloromethane	2.107	50	67709	51.446	ug/l	96
4) Vinyl Chloride	2.241	62	83507	55.729	ug/l	99
5) Bromomethane	2.625	94	56325	52.321	ug/l	100
6) Chloroethane	2.778	64	48122	55.006	ug/l	97
7) Trichlorofluoromethane	3.107	101	109548	54.758	ug/l	96
8) Diethyl Ether	3.515	74	38080	49.752	ug/l	86
9) 1,1,2-Trichlorotrifluo...	3.881	101	65624	48.414	ug/l	95
10) Methyl Iodide	4.076	142	85417	47.155	ug/l	91
11) Tert butyl alcohol	4.948	59	40892	200.871	ug/l #	92
12) 1,1-Dichloroethene	3.857	96	63369	50.404	ug/l #	80
13) Acrolein	3.717	56	20884	181.236	ug/l	99
14) Allyl chloride	4.466	41	108686	50.245	ug/l	97
15) Acrylonitrile	5.149	53	107386	254.981	ug/l	98
16) Acetone	3.936	43	86912	232.472	ug/l	94
17) Carbon Disulfide	4.180	76	179968	43.688	ug/l	98
18) Methyl Acetate	4.466	43	50604	46.100	ug/l	96
19) Methyl tert-butyl Ether	5.210	73	197772	51.974	ug/l	99
20) Methylene Chloride	4.698	84	83444	46.841	ug/l	89
21) trans-1,2-Dichloroethene	5.204	96	74464	50.549	ug/l	87
22) Diisopropyl ether	6.106	45	262542	57.090	ug/l	94
23) Vinyl Acetate	6.045	43	871707	272.823	ug/l	97
24) 1,1-Dichloroethane	6.003	63	131639	50.760	ug/l	98
25) 2-Butanone	6.972	43	156007	257.954	ug/l	95
26) 2,2-Dichloropropane	6.972	77	127297	49.803	ug/l	98
27) cis-1,2-Dichloroethene	6.978	96	87897	50.682	ug/l	91
28) Bromochloromethane	7.326	49	58663	53.820	ug/l	89
29) Tetrahydrofuran	7.338	42	106171	271.417	ug/l	94
30) Chloroform	7.496	83	140784	49.602	ug/l	99
31) Cyclohexane	7.771	56	122065	48.412	ug/l	91
32) 1,1,1-Trichloroethane	7.691	97	126812	50.677	ug/l	98
36) 1,1-Dichloropropene	7.905	75	108417	52.907	ug/l	97
37) Ethyl Acetate	7.064	43	71593	55.736	ug/l	97
38) Carbon Tetrachloride	7.893	117	115364	52.640	ug/l	95
39) Methylcyclohexane	9.179	83	129747	52.748	ug/l	91
40) Benzene	8.149	78	312864	51.993	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071322\
 Data File : VY009560.D
 Acq On : 13 Jul 2022 17:11
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By : John Carlone 07/14/2022
 Supervised By : Mahesh Dadoda 07/14/2022

Quant Time: Jul 14 02:45:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 08 11:00:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.295	41	40554m	60.801	ug/l	
42) 1,2-Dichloroethane	8.228	62	96532	53.434	ug/l	92
43) Isopropyl Acetate	8.265	43	132664	55.353	ug/l	96
44) Trichloroethene	8.929	130	89231	52.492	ug/l	100
45) 1,2-Dichloropropane	9.210	63	80034	52.556	ug/l	98
46) Dibromomethane	9.301	93	47898	53.590	ug/l	97
47) Bromodichloromethane	9.490	83	112026	52.660	ug/l	99
48) Methyl methacrylate	9.289	41	57359	55.548	ug/l	90
49) 1,4-Dioxane	9.283	88	14392	1093.415	ug/l	89
51) 4-Methyl-2-Pentanone	10.063	43	365816	290.238	ug/l	95
52) Toluene	10.240	92	211501	55.158	ug/l	95
53) t-1,3-Dichloropropene	10.459	75	123882	55.259	ug/l	100
54) cis-1,3-Dichloropropene	9.923	75	133632	53.458	ug/l	93
55) 1,1,2-Trichloroethane	10.642	97	70640	54.785	ug/l	94
56) Ethyl methacrylate	10.508	69	99708	58.245	ug/l	87
57) 1,3-Dichloropropane	10.782	76	115658	54.412	ug/l	96
58) 2-Chloroethyl Vinyl ether	9.776	63	225067	269.136	ug/l	96
59) 2-Hexanone	10.831	43	262981	301.110	ug/l	95
60) Dibromochloromethane	10.977	129	86700	55.527	ug/l	100
61) 1,2-Dibromoethane	11.081	107	67253	54.407	ug/l	99
64) Tetrachloroethene	10.715	164	87779	46.608	ug/l	97
65) Chlorobenzene	11.514	112	234360	52.974	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.587	131	88859	53.396	ug/l	98
67) Ethyl Benzene	11.587	91	414951	53.727	ug/l	97
68) m/p-Xylenes	11.697	106	335253	111.300	ug/l	92
69) o-Xylene	12.026	106	162601	56.318	ug/l	90
70) Styrene	12.038	104	280618	58.190	ug/l	95
71) Bromoform	12.203	173	60751	57.276	ug/l #	100
73) Isopropylbenzene	12.325	105	429692	50.686	ug/l	99
74) N-amyl acetate	12.142	43	134576	55.793	ug/l	93
75) 1,1,2,2-Tetrachloroethane	12.581	83	88957	52.407	ug/l	99
76) 1,2,3-Trichloropropane	12.630	75	62511m	47.638	ug/l	
77) Bromobenzene	12.605	156	103687	50.240	ug/l	95
78) n-propylbenzene	12.666	91	528827	51.633	ug/l	98
79) 2-Chlorotoluene	12.752	91	297455	51.875	ug/l	96
80) 1,3,5-Trimethylbenzene	12.806	105	363263	51.264	ug/l	96
81) trans-1,4-Dichloro-2-b...	12.374	75	33487	52.229	ug/l	89
82) 4-Chlorotoluene	12.849	91	313262	52.130	ug/l	97
83) tert-Butylbenzene	13.075	119	327344	51.924	ug/l	93
84) 1,2,4-Trimethylbenzene	13.117	105	369198	52.345	ug/l	97
85) sec-Butylbenzene	13.251	105	489460	52.717	ug/l	99
86) p-Isopropyltoluene	13.367	119	410942	53.098	ug/l	98
87) 1,3-Dichlorobenzene	13.361	146	211199	51.842	ug/l	99
88) 1,4-Dichlorobenzene	13.447	146	211824	51.374	ug/l	100
89) n-Butylbenzene	13.697	91	383456	53.058	ug/l	98
90) Hexachloroethane	13.959	117	77720	51.182	ug/l	88
91) 1,2-Dichlorobenzene	13.739	146	196199	52.565	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.355	75	15284	51.372	ug/l	83
93) 1,2,4-Trichlorobenzene	15.007	180	114281	50.201	ug/l	100
94) Hexachlorobutadiene	15.111	225	61068	48.793	ug/l	98
95) Naphthalene	15.233	128	240059	51.003	ug/l	98
96) 1,2,3-Trichlorobenzene	15.422	180	99084	50.253	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071322\
Data File : VY009560.D
Acq On : 13 Jul 2022 17:11
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC


Manual Integrations
APPROVED

Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022

Quant Time: Jul 14 02:45:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070722S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 08 11:00:56 2022
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.777 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion:168 Resp: 153684
Ion Ratio Lower Upper
168 100
99 50.5 46.9 70.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022



#2
Dichlorodifluoromethane
Concen: 141.603 ug/l
RT: 1.900 min Scan# 13
Delta R.T. -0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11


Tgt Ion: 85 Resp: 166886
Ion Ratio Lower Upper
85 100
87 32.0 16.1 48.2





#3
Chloromethane
Concen: 51.446 ug/l
RT: 2.107 min Scan# 47
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 50 Resp: 67709
Ion Ratio Lower Upper
50 100
52 33.1 24.6 37.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022




#4
Vinyl Chloride
Concen: 55.729 ug/l
RT: 2.241 min Scan# 69
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Tgt Ion: 62 Resp: 83507
Ion Ratio Lower Upper
62 100
64 30.6 25.0 37.6





#5
Bromomethane
Concen: 52.321 ug/l
RT: 2.625 min Scan# 11
Delta R.T. -0.012 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 94 Resp: 5632
Ion Ratio Lower Upper
94 100
96 91.9 73.5 110.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022




#6
Chloroethane
Concen: 55.006 ug/l
RT: 2.778 min Scan# 157
Delta R.T. -0.006 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Tgt Ion: 64 Resp: 48122
Ion Ratio Lower Upper
64 100
66 30.1 25.4 38.0





#7
Trichlorofluoromethane
Concen: 54.758 ug/l
RT: 3.107 min Scan# 211
Delta R.T. -0.006 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion:101 Resp: 10954
Ion Ratio Lower Upper
101 100
103 63.3 48.0 72.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022



#8
Diethyl Ether
Concen: 49.752 ug/l
RT: 3.515 min Scan# 278
Delta R.T. -0.006 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11

Tgt Ion: 74 Resp: 38080
Ion Ratio Lower Upper
74 100
45 95.7 55.1 165.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 48.414 ug/l
RT: 3.881 min Scan# 31
Delta R.T. -0.006 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion:101 Resp: 65624
Ion Ratio Lower Upper
101 100
85 45.8 40.6 60.8
151 85.0 64.9 97.3




Manual Integrations
APPROVED

Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022



#10
Methyl Iodide
Concen: 47.155 ug/l
RT: 4.076 min Scan# 370
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11


Tgt Ion:142 Resp: 85417
Ion Ratio Lower Upper
142 100
127 44.1 41.5 62.3
141 14.6 11.9 17.9





#11
Tert butyl alcohol
Concen: 200.871 ug/l
RT: 4.948 min Scan# 513
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 59 Resp: 40892
Ion Ratio Lower Upper
59 100
57 10.8 6.4 9.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022




#12
1,1-Dichloroethene
Concen: 50.404 ug/l
RT: 3.857 min Scan# 334
Delta R.T. -0.006 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11



Tgt Ion: 96 Resp: 63369
Ion Ratio Lower Upper
96 100
61 144.0 145.8 218.6#
98 63.6 52.5 78.7





#13

Acrolein

Concen: 181.236 ug/l

RT: 3.717 min Scan# 31

Delta R.T. -0.006 min

Lab File: VY009560.D

Acq: 13 Jul 2022 17:11

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 56 Resp: 20884

Ion Ratio Lower Upper

56 100





55 71.0 57.6 86.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 07/14/2022

Supervised By :Mahesh Dadoda 07/14/2022



#14

Allyl chloride

Concen: 50.245 ug/l

RT: 4.466 min Scan# 434

Delta R.T. 0.000 min

Lab File: VY009560.D

Acq: 13 Jul 2022 17:11




Tgt Ion: 41 Resp: 108686


Ion Ratio Lower Upper

41 100

39 73.3 58.8 88.2


76 36.5 25.1 37.7





#15
Acrylonitrile
Concen: 254.981 ug/l
RT: 5.149 min Scan# 546
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 53 Resp: 107380
Ion Ratio Lower Upper
53 100
52 81.3 67.0 100.4
51 36.4 29.8 44.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022




#16
Acetone
Concen: 232.472 ug/l
RT: 3.936 min Scan# 347
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Tgt Ion: 43 Resp: 86912
Ion Ratio Lower Upper
43 100
58 30.8 22.0 33.0





#17
Carbon Disulfide
Concen: 43.688 ug/l
RT: 4.180 min Scan# 387
Delta R.T. -0.006 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 76 Resp: 179968
Ion Ratio Lower Upper
76 100
78 8.4 7.4 11.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022



#18
Methyl Acetate
Concen: 46.100 ug/l
RT: 4.466 min Scan# 434
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11


Tgt Ion: 43 Resp: 50604
Ion Ratio Lower Upper
43 100
74 22.8 16.6 24.8





#19
Methyl tert-butyl Ether
Concen: 51.974 ug/l
RT: 5.210 min Scan# 519
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 73 Resp: 197772
Ion Ratio Lower Upper
73 100
57 22.6 18.6 27.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022




#20
Methylene Chloride
Concen: 46.841 ug/l
RT: 4.698 min Scan# 472
Delta R.T. -0.006 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Tgt Ion: 84 Resp: 83444
Ion Ratio Lower Upper
84 100
49 122.6 114.4 171.6
51 41.7 35.3 52.9
86 64.7 53.7 80.5





#21
trans-1,2-Dichloroethene
Concen: 50.549 ug/l
RT: 5.204 min Scan# 511
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 96 Resp: 74464
Ion Ratio Lower Upper
96 100
61 129.2 120.5 180.7
98 63.9 49.8 74.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022



#22
Diisopropyl ether
Concen: 57.090 ug/l
RT: 6.106 min Scan# 703
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11


Tgt Ion: 45 Resp: 262542
Ion Ratio Lower Upper
45 100
43 57.0 49.3 73.9
87 27.9 19.2 28.8
59 11.6 8.4 12.6





#23
 Vinyl Acetate
 Concen: 272.823 ug/l
 RT: 6.045 min Scan# 694
 Delta R.T. -0.006 min
 Lab File: VY009560.D
 Acq: 13 Jul 2022 17:11




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC




Tgt Ion: 43 Resp: 87170
 Ion Ratio Lower Upper
 43 100
 86 10.0 7.1 10.7

Manual Integrations
 APPROVED



Reviewed By :John Carlone 07/14/2022
 Supervised By :Mahesh Dadoda 07/14/2022




#24
 1,1-Dichloroethane
 Concen: 50.760 ug/l
 RT: 6.003 min Scan# 686
 Delta R.T. -0.006 min
 Lab File: VY009560.D
 Acq: 13 Jul 2022 17:11




Tgt Ion: 63 Resp: 131639
 Ion Ratio Lower Upper
 63 100
 98 7.6 3.6 10.8
 100 4.6 2.0 5.8





#25
2-Butanone
Concen: 257.954 ug/l
RT: 6.972 min Scan# 845
Delta R.T. -0.006 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 43 Resp: 15600
Ion Ratio Lower Upper
43 100
72 24.7 17.7 26.5

Manual Integrations
APPROVED



Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022




#26
2,2-Dichloropropane
Concen: 49.803 ug/l
RT: 6.972 min Scan# 845
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11



Tgt Ion: 77 Resp: 127297
Ion Ratio Lower Upper
77 100
97 21.6 10.4 31.4





#27

cis-1,2-Dichloroethene

Concen: 50.682 ug/l

RT: 6.978 min Scan# 846

Delta R.T. 0.000 min

Lab File: VY009560.D


Acq: 13 Jul 2022 17:11

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC






Tgt Ion: 96	Resp: 8789
Ion Ratio	Lower Upper
96	100
61	144.6 0.0 319.8
98	64.5 0.0 127.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 07/14/2022

Supervised By :Mahesh Dadoda 07/14/2022



#28

Bromochloromethane

Concen: 53.820 ug/l




RT: 7.326 min Scan# 903


Delta R.T. -0.000 min

Lab File: VY009560.D

Acq: 13 Jul 2022 17:11

Tgt Ion: 49	Resp: 58663
Ion Ratio	Lower Upper
49	100
129	2.0 0.0 2.2
130	78.3 55.4 83.0





#29

Tetrahydrofuran

Concen: 271.417 ug/l

RT: 7.338 min Scan# 905

Delta R.T. -0.000 min

Lab File: VY009560.D


Acq: 13 Jul 2022 17:11

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC





Tgt Ion:	42	Resp:	10617
Ion Ratio	Lower	Upper	
42	100		
72	41.8	30.3	45.5
71	39.8	29.0	43.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 07/14/2022

Supervised By :Mahesh Dadoda 07/14/2022



#30

Chloroform

Concen: 49.602 ug/l




RT: 7.496 min Scan# 931


Delta R.T. 0.000 min

Lab File: VY009560.D

Acq: 13 Jul 2022 17:11


Tgt Ion:	83	Resp:	140784
Ion Ratio	Lower	Upper	
83	100		
85	65.5	51.8	77.6





#31
Cyclohexane
Concen: 48.412 ug/l
RT: 7.771 min Scan# 911
Delta R.T. -0.006 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 56 Resp: 122065
Ion Ratio Lower Upper
56 100
69 31.5 26.7 40.1
84 93.8 67.2 100.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022




#32
1,1,1-Trichloroethane
Concen: 50.677 ug/l
RT: 7.691 min Scan# 963
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11



Tgt Ion: 97 Resp: 126812
Ion Ratio Lower Upper
97 100
99 63.6 50.7 76.1
61 43.4 37.6 56.4









#33
1,2-Dichloroethane-d4
Concen: 48.683 ug/l
RT: 8.130 min Scan# 1036
Delta R.T. -0.006 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 65 Resp: 5288
Ion Ratio Lower Upper
65 100
67 54.7 0.0 99.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.679 min Scan# 1125
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11


Tgt Ion:114 Resp: 239217
Ion Ratio Lower Upper
114 100
63 19.3 0.0 43.2
88 14.6 0.0 31.8





#35
Dibromofluoromethane
Concen: 45.701 ug/l
RT: 7.710 min Scan# 90
Delta R.T. -0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:113 Resp: 5752
Ion Ratio Lower Upper
113 100
111 103.2 82.6 124.0
192 20.5 17.6 26.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022




#36
1,1-Dichloropropene
Concen: 52.907 ug/l
RT: 7.905 min Scan# 998
Delta R.T. -0.006 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Tgt Ion: 75 Resp: 108417
Ion Ratio Lower Upper
75 100
110 37.0 17.1 51.2
77 30.6 24.4 36.6





#37
Ethyl Acetate
Concen: 55.736 ug/l
RT: 7.064 min Scan# 860
Delta R.T. 0.001 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 7159
Ion Ratio Lower Upper
43 100
61 13.1 9.7 14.5
70 10.4 7.1 10.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022



#38
Carbon Tetrachloride
Concen: 52.640 ug/l
RT: 7.893 min Scan# 996
Delta R.T. 0.001 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11

Tgt Ion:117 Resp: 115364
Ion Ratio Lower Upper
117 100
119 92.0 77.9 116.9
121 32.6 24.8 37.2





#39

Methylcyclohexane

Concen: 52.748 ug/l

RT: 9.179 min Scan# 11

Delta R.T. 0.000 min

Lab File: VY009560.D

Acq: 13 Jul 2022 17:11

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 83 Resp: 12974





Ion	Ratio	Lower	Upper
83	100		
55	74.1	66.5	99.7
98	52.0	37.9	56.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 07/14/2022

Supervised By :Mahesh Dadoda 07/14/2022



#40

Benzene

Concen: 51.993 ug/l

RT: 8.149 min Scan# 1038




Delta R.T. -0.006 min


Lab File: VY009560.D

Acq: 13 Jul 2022 17:11

Tgt Ion: 78 Resp: 312864


Ion	Ratio	Lower	Upper
78	100		
77	23.2	20.4	30.6





#41
Methacrylonitrile
Concen: 60.801 ug/l m
RT: 7.295 min Scan# 898
Delta R.T. -0.012 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 41 Resp: 40554
Ion Ratio Lower Upper
41 100
39 0.0 92.9 139.3
67 0.0 0.0 0.0
52 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022



#42
1,2-Dichloroethane
Concen: 53.434 ug/l
RT: 8.228 min Scan# 1051
Delta R.T. -0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11


Tgt Ion: 62 Resp: 96532
Ion Ratio Lower Upper
62 100
98 10.1 0.0 14.8





#43
Isopropyl Acetate
Concen: 55.353 ug/l
RT: 8.265 min Scan# 1105
Delta R.T. 0.001 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 132664
Ion Ratio Lower Upper
43 100
61 27.4 20.2 30.2
87 12.9 8.9 13.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022



#44
Trichloroethene
Concen: 52.492 ug/l
RT: 8.929 min Scan# 1166
Delta R.T. -0.006 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11


Tgt Ion:130 Resp: 89231
Ion Ratio Lower Upper
130 100
95 94.6 0.0 188.2





#45
1,2-Dichloropropane
Concen: 52.556 ug/l
RT: 9.210 min Scan# 1212
Delta R.T. 0.001 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 63 Resp: 80034
Ion Ratio Lower Upper
63 100
65 29.9 23.0 34.6




Manual Integrations
APPROVED

Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022



#46
Dibromomethane
Concen: 53.590 ug/l
RT: 9.301 min Scan# 1227
Delta R.T. -0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11


Tgt Ion: 93 Resp: 47898
Ion Ratio Lower Upper
93 100
95 83.6 64.3 96.5
174 107.3 83.4 125.0





#47
Bromodichloromethane
Concen: 52.660 ug/l
RT: 9.490 min Scan# 1258
Delta R.T. -0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 83 Resp: 112020
Ion Ratio Lower Upper
83 100
85 65.3 51.9 77.9
127 9.0 6.6 10.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022




#48
Methyl methacrylate
Concen: 55.548 ug/l
RT: 9.289 min Scan# 1225
Delta R.T. -0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11



Tgt Ion: 41 Resp: 57359
Ion Ratio Lower Upper
41 100
69 85.5 58.2 87.2
39 62.5 47.4 71.0





#49

1,4-Dioxane

Concen: 1093.415 ug/l

RT: 9.283 min Scan# 11

Delta R.T. -0.006 min

Lab File: VY009560.D

Acq: 13 Jul 2022 17:11

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC





Tgt Ion:	88	Resp:	1439
Ion Ratio	Lower	Upper	
88	100		
43	33.6	32.8	49.2
58	65.4	59.8	89.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 07/14/2022

Supervised By :Mahesh Dadoda 07/14/2022



#50

Toluene-d8

Concen: 49.317 ug/l




RT: 10.173 min Scan# 1370


Delta R.T. -0.000 min

Lab File: VY009560.D

Acq: 13 Jul 2022 17:11


Tgt Ion:	98	Resp:	170576
Ion Ratio	Lower	Upper	
98	100		
100	65.3	50.2	75.2





#51
4-Methyl-2-Pentanone
Concen: 290.238 ug/l
RT: 10.063 min Scan# 1352
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 43 Resp: 365810
Ion Ratio Lower Upper
43 100
58 38.1 28.1 42.1

Manual Integrations
APPROVED



Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022




#52
Toluene
Concen: 55.158 ug/l
RT: 10.240 min Scan# 1381
Delta R.T. -0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Tgt Ion: 92 Resp: 211501
Ion Ratio Lower Upper
92 100
91 171.2 142.6 213.8





#53
t-1,3-Dichloropropene
Concen: 55.259 ug/l
RT: 10.459 min Scan# 1417
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 75 Resp: 123882
Ion Ratio Lower Upper
75 100
77 32.0 25.4 38.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022




#54
cis-1,3-Dichloropropene
Concen: 53.458 ug/l
RT: 9.923 min Scan# 1329
Delta R.T. -0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Tgt Ion: 75 Resp: 133632
Ion Ratio Lower Upper
75 100
77 32.8 25.6 38.4
39 51.7 47.6 71.4





#55
1,1,2-Trichloroethane
Concen: 54.785 ug/l
RT: 10.642 min Scan# 1447
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 97 Resp: 70640


Ion	Ratio	Lower	Upper
97	100		
83	79.1	69.8	104.8
85	52.2	44.3	66.5
99	59.6	49.8	74.6

Manual Integrations
APPROVED

Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022






#56
Ethyl methacrylate
Concen: 58.245 ug/l
RT: 10.508 min Scan# 1425
Delta R.T. 0.006 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11



Tgt Ion: 69 Resp: 99708

Ion	Ratio	Lower	Upper
69	100		
41	67.6	63.4	95.0
39	37.1	35.8	53.8





#57

1,3-Dichloropropane

Concen: 54.412 ug/l

RT: 10.782 min Scan# 1470

Delta R.T. 0.000 min

Lab File: VY009560.D

Acq: 13 Jul 2022 17:11

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 76 Resp: 115658

Ion Ratio Lower Upper

76 100





78 33.6 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 07/14/2022

Supervised By :Mahesh Dadoda 07/14/2022



#58

2-Chloroethyl Vinyl ether

Concen: 269.136 ug/l

RT: 9.776 min Scan# 1305

Delta R.T. 0.000 min

Lab File: VY009560.D




Acq: 13 Jul 2022 17:11


Tgt Ion: 63 Resp: 225067

Ion Ratio Lower Upper

63 100


106 28.9 21.5 32.3





#59
2-Hexanone
Concen: 301.110 ug/l
RT: 10.831 min Scan# 1477
Delta R.T. 0.006 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 26298
Ion Ratio Lower Upper
43 100
58 52.6 24.4 73.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022



#60
Dibromochloromethane
Concen: 55.527 ug/l
RT: 10.977 min Scan# 1502
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11


Tgt Ion:129 Resp: 86700
Ion Ratio Lower Upper
129 100
127 77.3 38.7 116.1





#61
1,2-Dibromoethane
Concen: 54.407 ug/l
RT: 11.081 min Scan# 1519
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:107 Resp: 6725
Ion Ratio Lower Upper
107 100
109 95.4 75.4 113.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022




#62
4-Bromofluorobenzene
Concen: 54.508 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Tgt Ion: 95 Resp: 96846
Ion Ratio Lower Upper
95 100
174 83.3 0.0 177.0
176 83.3 0.0 175.8





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.490 min Scan# 11586
Delta R.T. 0.001 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion:117 Resp: 225250
Ion Ratio Lower Upper
117 100
82 52.5 43.0 64.4
119 32.6 25.4 38.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022



#64
Tetrachloroethene
Concen: 46.608 ug/l
RT: 10.715 min Scan# 1459
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11


Tgt Ion:164 Resp: 87779
Ion Ratio Lower Upper
164 100
166 128.7 100.4 150.6
129 86.4 71.4 107.0
131 86.3 70.8 106.2





#65
Chlorobenzene
Concen: 52.974 ug/l
RT: 11.514 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion:112 Resp: 234360
Ion Ratio Lower Upper
112 100
114 31.7 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022



#66
1,1,1,2-Tetrachloroethane
Concen: 53.396 ug/l
RT: 11.587 min Scan# 1602
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11


Tgt Ion:131 Resp: 88859
Ion Ratio Lower Upper
131 100
133 95.0 47.9 143.6
119 65.0 30.8 92.3





#67
Ethyl Benzene
Concen: 53.727 ug/l
RT: 11.587 min Scan# 1602
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 91 Resp: 41495
Ion Ratio Lower Upper
91 100
106 32.1 24.5 36.7

Manual Integrations
APPROVED



Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022




#68
m/p-Xylenes
Concen: 111.300 ug/l
RT: 11.697 min Scan# 1620
Delta R.T. -0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Tgt Ion:106 Resp: 335253
Ion Ratio Lower Upper
106 100
91 195.4 166.6 249.8





#69
o-Xylene
Concen: 56.318 ug/l
RT: 12.026 min Scan# 1674
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion:106 Resp: 16260
Ion Ratio Lower Upper
106 100
91 204.1 110.3 331.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022



#70
Styrene
Concen: 58.190 ug/l
RT: 12.038 min Scan# 1676
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11


Tgt Ion:104 Resp: 280618
Ion Ratio Lower Upper
104 100
78 47.8 42.4 63.6
103 53.7 44.4 66.6





#71
Bromoform
Concen: 57.276 ug/l
RT: 12.203 min Scan# 1703
Delta R.T. -0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:173 Resp: 6075
Ion Ratio Lower Upper
173 100
175 49.5 24.6 74.0
254 0.2 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.422 min Scan# 1903
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Tgt Ion:152 Resp: 123960
Ion Ratio Lower Upper
152 100
115 56.6 28.2 84.7
150 176.0 0.0 347.6





#73
Isopropylbenzene
Concen: 50.686 ug/l
RT: 12.325 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion:105 Resp: 429692
Ion Ratio Lower Upper
105 100
120 27.1 13.4 40.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022



#74
N-ethyl acetate
Concen: 55.793 ug/l
RT: 12.142 min Scan# 1693
Delta R.T. -0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11

Tgt Ion: 43 Resp: 134576
Ion Ratio Lower Upper
43 100
70 41.8 28.7 43.1
55 22.7 17.3 25.9
61 23.8 17.0 25.6









#75
1,1,2,2-Tetrachloroethane
Concen: 52.407 ug/l
RT: 12.581 min Scan# 1765
Delta R.T. 0.006 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 83 Resp: 8895
Ion Ratio Lower Upper
83 100
131 10.8 5.6 16.8
85 63.5 32.3 96.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022



#76
1,2,3-Trichloropropane
Concen: 47.638 ug/l m
RT: 12.630 min Scan# 1773
Delta R.T. 0.001 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11

Tgt Ion: 75 Resp: 62511
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77

Bromobenzene

Concen: 50.240 ug/l

RT: 12.605 min Scan# 1769

Delta R.T. 0.000 min

Lab File: VY009560.D

Acq: 13 Jul 2022 17:11

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC





Tgt Ion:	156	Resp:	10368
Ion Ratio	Lower	Upper	
156	100		
77	180.6	95.5	286.4
158	96.4	48.0	144.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 07/14/2022

Supervised By :Mahesh Dadoda 07/14/2022



#78

n-propylbenzene

Concen: 51.633 ug/l




RT: 12.666 min Scan# 1779


Delta R.T. 0.000 min

Lab File: VY009560.D

Acq: 13 Jul 2022 17:11

Tgt Ion:	91	Resp:	528827
Ion Ratio	Lower	Upper	
91	100		
120	23.6	11.3	33.9









#79
2-Chlorotoluene
Concen: 51.875 ug/l
RT: 12.752 min Scan# 1793
Delta R.T. 0.001 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 91 Resp: 297458
Ion Ratio Lower Upper
91 100
126 35.7 16.9 50.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022



#80
1,3,5-Trimethylbenzene
Concen: 51.264 ug/l
RT: 12.806 min Scan# 1802
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11


Tgt Ion:105 Resp: 363263
Ion Ratio Lower Upper
105 100
120 51.4 24.4 73.4





#81
trans-1,4-Dichloro-2-butene
Concen: 52.229 ug/l
RT: 12.374 min Scan# 1731
Delta R.T. 0.001 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 75 Resp: 3348
Ion Ratio Lower Upper
75 100
53 90.9 85.3 127.9
89 46.2 37.0 55.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022




#82
4-Chlorotoluene
Concen: 52.130 ug/l
RT: 12.849 min Scan# 1809
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Tgt Ion: 91 Resp: 313262
Ion Ratio Lower Upper
91 100
126 36.0 17.0 51.0





#83
tert-Butylbenzene
Concen: 51.924 ug/l
RT: 13.075 min Scan# 1845
Delta R.T. 0.007 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:119 Resp: 327344
Ion Ratio Lower Upper
119 100
91 62.5 34.8 104.4
134 27.1 13.3 39.9

Manual Integrations
APPROVED



Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022




#84
1,2,4-Trimethylbenzene
Concen: 52.345 ug/l
RT: 13.117 min Scan# 1853
Delta R.T. 0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Tgt Ion:105 Resp: 369198
Ion Ratio Lower Upper
105 100
120 46.4 22.3 66.9





#85
 sec-Butylbenzene
 Concen: 52.717 ug/l
 RT: 13.251 min Scan# 1875
 Delta R.T. 0.000 min
 Lab File: VY009560.D
 Acq: 13 Jul 2022 17:11




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC



Tgt Ion:105 Resp: 489460
 Ion Ratio Lower Upper
 105 100
 134 20.9 10.2 30.4




Manual Integrations
 APPROVED


Reviewed By :John Carlone 07/14/2022
 Supervised By :Mahesh Dadoda 07/14/2022



#86
 p-Isopropyltoluene
 Concen: 53.098 ug/l
 RT: 13.367 min Scan# 1894
 Delta R.T. 0.000 min
 Lab File: VY009560.D
 Acq: 13 Jul 2022 17:11


Tgt Ion:119 Resp: 410942
 Ion Ratio Lower Upper
 119 100
 134 26.9 13.2 39.5
 91 23.1 12.4 37.2





#87
 1,3-Dichlorobenzene
 Concen: 51.842 ug/l
 RT: 13.361 min Scan# 1893
 Delta R.T. 0.000 min
 Lab File: VY009560.D
 Acq: 13 Jul 2022 17:11




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC




Tgt Ion:146 Resp: 211199
 Ion Ratio Lower Upper
 146 100
 111 38.5 19.5 58.5
 148 64.4 31.6 95.0

Manual Integrations
 APPROVED



Reviewed By :John Carlone 07/14/2022
 Supervised By :Mahesh Dadoda 07/14/2022



#88
 1,4-Dichlorobenzene
 Concen: 51.374 ug/l
 RT: 13.447 min Scan# 1907
 Delta R.T. 0.007 min
 Lab File: VY009560.D
 Acq: 13 Jul 2022 17:11




Tgt Ion:146 Resp: 211824
 Ion Ratio Lower Upper
 146 100
 111 38.3 19.3 57.8
 148 65.1 32.6 97.8





#89
n-Butylbenzene
Concen: 53.058 ug/l
RT: 13.697 min Scan# 1947
Delta R.T. 0.007 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 91 Resp: 383450
Ion Ratio Lower Upper
91 100
92 54.3 26.3 78.9
134 27.3 13.0 38.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022



#90
Hexachloroethane
Concen: 51.182 ug/l
RT: 13.959 min Scan# 1991
Delta R.T. 0.001 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11

Tgt Ion:117 Resp: 77720
Ion Ratio Lower Upper
117 100
201 75.3 43.0 128.8









#91
1,2-Dichlorobenzene
Concen: 52.565 ug/l
RT: 13.739 min Scan# 1954
Delta R.T. 0.006 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion:146 Resp: 196199
Ion Ratio Lower Upper
146 100
111 39.0 20.3 60.9
148 63.1 32.3 96.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022



#92
1,2-Dibromo-3-Chloropropane
Concen: 51.372 ug/l
RT: 14.355 min Scan# 2056
Delta R.T. -0.000 min
Lab File: VY009560.D
Acq: 13 Jul 2022 17:11

Tgt Ion: 75 Resp: 15284
Ion Ratio Lower Upper
75 100
155 86.2 38.0 113.9
157 116.3 47.8 143.3





#93

1,2,4-Trichlorobenzene

Concen: 50.201 ug/l

RT: 15.007 min Scan# 2162

Delta R.T. 0.006 min

Lab File: VY009560.D

Acq: 13 Jul 2022 17:11

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion:180 Resp: 11428

Ion Ratio Lower Upper

180 100

182 95.1 47.3 141.9





145 29.2 14.6 43.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 07/14/2022

Supervised By :Mahesh Dadoda 07/14/2022



#94

Hexachlorobutadiene

Concen: 48.793 ug/l

RT: 15.111 min Scan# 2180

Delta R.T. 0.006 min

Lab File: VY009560.D

Acq: 13 Jul 2022 17:11




Tgt Ion:225 Resp: 61068


Ion Ratio Lower Upper

225 100

223 63.2 30.8 92.3

227 64.0 31.6 94.7





#95

Naphthalene

Concen: 51.003 ug/l

RT: 15.233 min Scan# 2200

Delta R.T. -0.000 min

Lab File: VY009560.D

Acq: 13 Jul 2022 17:11

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC





Tgt Ion:	128	Resp:	240059
Ion Ratio	Lower	Upper	
128	100		
127	13.0	11.0	16.4
129	11.0	8.4	12.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 07/14/2022

Supervised By :Mahesh Dadoda 07/14/2022



#96

1,2,3-Trichlorobenzene

Concen: 50.253 ug/l

RT: 15.422 min Scan# 2231

Delta R.T. -0.000 min

Lab File: VY009560.D

Acq: 13 Jul 2022 17:11

Tgt Ion:	180	Resp:	99084
Ion Ratio	Lower	Upper	
180	100		
182	94.6	47.3	141.9
145	30.8	15.3	45.9

