



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071322\
 Data File : VY009547.D
 Acq On : 13 Jul 2022 08:42
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070722S.M
 Title : SW846 8260
 Last Update : Fri Jul 08 11:00:56 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1739

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.9	16774	PASS
75	95	30	60	52.5	42056	PASS
95	95	100	100	100.0	80155	PASS
96	95	5	9	7.0	5589	PASS
173	174	0.00	2	1.3	986	PASS
174	95	50	100	91.3	73221	PASS
175	174	5	9	7.5	5510	PASS
176	174	95	101	96.9	70952	PASS
177	176	5	9	6.4	4560	PASS