

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071422\
 Data File : VY009563.D
 Acq On : 14 Jul 2022 14:38
 Operator : KP/MD
 Sample : VY0714SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0714SBL01

Quant Time: Jul 14 14:55:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 08 11:00:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	109036	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	179716	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	165277	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	73717	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	42559	56.216	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.440%
35) Dibromofluoromethane	7.716	113	39557	41.832	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	83.660%
50) Toluene-d8	10.173	98	145634	57.369	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	114.740%
62) 4-Bromofluorobenzene	12.477	95	45158	33.831	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	67.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071422\
Data File : VY009563.D
Acq On : 14 Jul 2022 14:38
Operator : KP/MD
Sample : VY0714SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0714SBL01

Quant Time: Jul 14 14:55:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070722S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 08 11:00:56 2022
Response via : Initial Calibration

