

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071423\
 Data File : VY014645.D
 Acq On : 14 Jul 2023 14:16
 Operator : SY/MD
 Sample : 03615-04
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ-210

Quant Time: Jul 14 23:10:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	233031	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	400193	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	373325	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	180150	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	142464	53.649	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.300%
35) Dibromofluoromethane	7.716	113	124839	49.004	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.000%
50) Toluene-d8	10.173	98	473435	50.407	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.820%
62) 4-Bromofluorobenzene	12.477	95	182377	50.642	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	101.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071423\
Data File : VY014645.D
Acq On : 14 Jul 2023 14:16
Operator : SY/MD
Sample : 03615-04
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-210

Quant Time: Jul 14 23:10:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

