

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071423\
 Data File : VY014655.D
 Acq On : 14 Jul 2023 18:15
 Operator : SY/MD
 Sample : 03632-05
 Misc : 7.84g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-05-(6-8)

Quant Time: Jul 14 23:14:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	223173	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	385871	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	359598	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	173987	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	138556	54.482	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 108.960%	
35) Dibromofluoromethane	7.710	113	120631	49.110	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 98.220%	
50) Toluene-d8	10.179	98	456230	50.378	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 100.760%	
62) 4-Bromofluorobenzene	12.477	95	174859	50.356	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 100.720%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.704	84	29809	7.824	ug/l	90
39) Methylcyclohexane	9.179	83	5164	1.079	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071423\
Data File : VY014655.D
Acq On : 14 Jul 2023 18:15
Operator : SY/MD
Sample : 03632-05
Misc : 7.84g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-05-(6-8)

Quant Time: Jul 14 23:14:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

