

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071423\
 Data File : VY014661.D
 Acq On : 14 Jul 2023 20:36
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 14 23:18:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	85	0.00
2 T	Dichlorodifluoromethane	0.407	0.394	3.2	86	0.00
3 P	Chloromethane	0.458	0.504	-10.0	95	0.00
4 C	Vinyl Chloride	0.512	0.554	-8.2#	91	0.00
5 T	Bromomethane	0.338	0.389	-15.1	99	0.00
6 T	Chloroethane	0.320	0.364	-13.7	95	0.00
7 T	Trichlorofluoromethane	0.860	0.909	-5.7	88	0.00
8 T	Diethyl Ether	0.320	0.360	-12.5	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.527	0.542	-2.8	85	0.00
10 T	Methyl Iodide	0.493	0.604	-22.5	95	0.00
11 T	Tert butyl alcohol	0.125	0.106	15.2	81	0.00
12 CM	1,1-Dichloroethene	0.492	0.513	-4.3#	86	0.00
13 T	Acrolein	0.060	0.055	8.3	86	0.00
14 T	Allyl chloride	0.874	0.929	-6.3	89	0.00
15 T	Acrylonitrile	0.176	0.215	-22.2	102	0.00
16 T	Acetone	0.181	0.173	4.4	76	0.00
17 T	Carbon Disulfide	1.410	1.409	0.1	86	0.00
18 T	Methyl Acetate	0.693	0.968	-39.7#	117	0.00
19 T	Methyl tert-butyl Ether	1.599	1.868	-16.8	97	0.00
20 T	Methylene Chloride	0.656	0.632	3.7	89	0.00
21 T	trans-1,2-Dichloroethene	0.559	0.597	-6.8	89	0.00
22 T	Diisopropyl ether	1.777	2.014	-13.3	93	0.00
23 T	Vinyl Acetate	1.064	1.160	-9.0	90	0.00
24 P	1,1-Dichloroethane	1.017	1.134	-11.5	92	0.00
25 T	2-Butanone	0.264	0.299	-13.3	94	0.00
26 T	2,2-Dichloropropane	0.969	0.852	12.1	73	0.00
27 T	cis-1,2-Dichloroethene	0.650	0.725	-11.5	92	0.00
28 T	Bromochloromethane	0.438	0.468	-6.8	95	0.00
29 T	Tetrahydrofuran	0.163	0.201	-23.3	105	0.00
30 C	Chloroform	1.049	1.183	-12.8#	94	0.00
31 T	Cyclohexane	0.946	0.916	3.2	85	0.00
32 T	1,1,1-Trichloroethane	0.944	1.033	-9.4	91	0.00
33 S	1,2-Dichloroethane-d4	0.570	0.632	-10.9	96	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	88	0.00
35 S	Dibromofluoromethane	0.318	0.335	-5.3	91	0.00
36 T	1,1-Dichloropropene	0.476	0.495	-4.0	88	0.00
37 T	Ethyl Acetate	0.328	0.378	-15.2	103	0.00
38 T	Carbon Tetrachloride	0.503	0.537	-6.8	90	0.00
39 T	Methylcyclohexane	0.620	0.613	1.1	84	0.00
40 TM	Benzene	1.354	1.454	-7.4	91	0.00
41 T	Methacrylonitrile	0.169	0.180	-6.5	94	0.00
42 TM	1,2-Dichloroethane	0.429	0.495	-15.4	97	0.00
43 T	Isopropyl Acetate	0.624	0.745	-19.4	102	0.00
44 TM	Trichloroethene	0.379	0.406	-7.1	91	0.00
45 C	1,2-Dichloropropane	0.353	0.388	-9.9#	93	0.00
46 T	Dibromomethane	0.214	0.239	-11.7	94	0.00
47 T	Bromodichloromethane	0.505	0.567	-12.3	94	0.00
48 T	Methyl methacrylate	0.281	0.327	-16.4	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071423\
 Data File : VY014661.D
 Acq On : 14 Jul 2023 20:36
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 14 23:18:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.004	-33.3#	103	0.00
50 S	Toluene-d8	1.173	1.202	-2.5	90	0.00
51 T	4-Methyl-2-Pentanone	0.328	0.394	-20.1	103	0.00
52 CM	Toluene	0.866	0.928	-7.2#	91	0.00
53 T	t-1,3-Dichloropropene	0.553	0.589	-6.5	90	0.00
54 T	cis-1,3-Dichloropropene	0.601	0.629	-4.7	88	0.00
55 T	1,1,2-Trichloroethane	0.299	0.343	-14.7	97	0.00
56 T	Ethyl methacrylate	0.456	0.519	-13.8	95	0.00
57 T	1,3-Dichloropropane	0.515	0.580	-12.6	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.213	0.227	-6.6	93	0.00
59 T	2-Hexanone	0.239	0.285	-19.2	101	0.00
60 T	Dibromochloromethane	0.366	0.407	-11.2	93	0.00
61 T	1,2-Dibromoethane	0.298	0.338	-13.4	95	0.00
62 S	4-Bromofluorobenzene	0.450	0.462	-2.7	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	89	0.00
64 T	Tetrachloroethene	0.402	0.439	-9.2	95	0.00
65 PM	Chlorobenzene	1.065	1.131	-6.2	92	0.00
66 T	1,1,1,2-Tetrachloroethane	0.402	0.438	-9.0	93	0.00
67 C	Ethyl Benzene	1.924	2.050	-6.5#	91	0.00
68 T	m/p-Xylenes	0.734	0.779	-6.1	91	0.00
69 T	o-Xylene	0.715	0.764	-6.9	92	0.00
70 T	Styrene	1.206	1.298	-7.6	91	0.00
71 P	Bromoform	0.272	0.314	-15.4	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
73 T	Isopropylbenzene	3.915	4.027	-2.9	90	0.00
74 T	N-amyl acetate	1.311	1.430	-9.1	95	0.00
75 P	1,1,2,2-Tetrachloroethane	0.893	0.977	-9.4	97	0.00
76 T	1,2,3-Trichloropropane	0.631	0.727	-15.2	102	0.00
77 T	Bromobenzene	0.894	0.961	-7.5	93	0.00
78 T	n-propylbenzene	4.718	4.916	-4.2	90	0.00
79 T	2-Chlorotoluene	2.717	2.863	-5.4	91	0.00
80 T	1,3,5-Trimethylbenzene	3.270	3.492	-6.8	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.326	0.329	-0.9	87	0.00
82 T	4-Chlorotoluene	2.799	2.922	-4.4	91	0.00
83 T	tert-Butylbenzene	2.900	2.994	-3.2	89	0.00
84 T	1,2,4-Trimethylbenzene	3.209	3.774	-17.6	102	0.00
85 T	sec-Butylbenzene	4.255	4.358	-2.4	89	0.00
86 T	p-Isopropyltoluene	3.474	3.615	-4.1	89	0.00
87 T	1,3-Dichlorobenzene	1.739	1.861	-7.0	92	0.00
88 T	1,4-Dichlorobenzene	1.743	1.862	-6.8	92	0.00
89 T	n-Butylbenzene	3.380	3.538	-4.7	90	0.00
90 T	Hexachloroethane	0.714	0.749	-4.9	91	0.00
91 T	1,2-Dichlorobenzene	1.594	1.717	-7.7	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.171	0.204	-19.3	106	0.00
93 T	1,2,4-Trichlorobenzene	1.012	1.099	-8.6	91	0.00
94 T	Hexachlorobutadiene	0.556	0.597	-7.4	90	0.00
95 T	Naphthalene	2.357	3.051	-29.4#	110	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071423\
Data File : VY014661.D
Acq On : 14 Jul 2023 20:36
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jul 14 23:18:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.899	1.007	-12.0	95	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6