

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071524\
 Data File : VY018819.D
 Acq On : 15 Jul 2024 17:49
 Operator : SY/MD
 Sample : P3208-05RE
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CONC-PADSRE

Quant Time: Jul 16 01:28:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 09 14:28:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	156889	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.679	114	279563	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	259707	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	111234	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	80878	52.388	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.780%
35) Dibromofluoromethane	7.704	113	31846	18.188	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	36.380%#
50) Toluene-d8	10.173	98	319924	48.227	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.460%
62) 4-Bromofluorobenzene	12.477	95	110562	48.804	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	97.600%
Target Compounds						
						Qvalue
16) Acetone	3.942	43	3557	12.497	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071524\
Data File : VY018819.D
Acq On : 15 Jul 2024 17:49
Operator : SY/MD
Sample : P3208-05RE
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CONC-PADSRE

Quant Time: Jul 16 01:28:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 09 14:28:51 2024
Response via : Initial Calibration

