

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071524\
 Data File : VY018810.D
 Acq On : 15 Jul 2024 11:08
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 15 13:12:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 09 14:28:51 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	84	0.00
2 T	Dichlorodifluoromethane	0.419	0.390	6.9	73	0.00
3 P	Chloromethane	0.637	0.648	-1.7	84	0.00
4 C	Vinyl Chloride	0.745	0.729	2.1#	79	0.00
5 T	Bromomethane	0.512	0.500	2.3	87	0.00
6 T	Chloroethane	0.484	0.475	1.9	83	0.00
7 T	Trichlorofluoromethane	0.973	0.973	0.0	79	0.00
8 T	Diethyl Ether	0.237	0.260	-9.7	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.499	0.472	5.4	75	0.00
10 T	Methyl Iodide	0.538	0.581	-8.0	82	0.00
11 T	Tert butyl alcohol	0.050	0.039	22.0	88	0.00
12 CM	1,1-Dichloroethene	0.483	0.471	2.5#	80	0.00
13 T	Acrolein	0.040	0.034	15.0	77	0.00
14 T	Allyl chloride	0.598	0.622	-4.0	86	0.00
15 T	Acrylonitrile	0.094	0.102	-8.5	90	0.00
16 T	Acetone	0.091	0.110	-20.9	104	-0.01
17 T	Carbon Disulfide	1.353	1.356	-0.2	80	0.00
18 T	Methyl Acetate	0.264	0.253	4.2	87	0.00
19 T	Methyl tert-butyl Ether	1.121	1.211	-8.0	88	0.00
20 T	Methylene Chloride	0.610	0.582	4.6	92	0.00
21 T	trans-1,2-Dichloroethene	0.510	0.528	-3.5	84	0.00
22 T	Diisopropyl ether	1.302	1.354	-4.0	88	0.00
23 T	Vinyl Acetate	1.026	1.111	-8.3	88	0.00
24 P	1,1-Dichloroethane	0.831	0.855	-2.9	87	0.00
25 T	2-Butanone	0.127	0.141	-11.0	94	0.00
26 T	2,2-Dichloropropane	0.771	0.789	-2.3	85	0.00
27 T	cis-1,2-Dichloroethene	0.594	0.612	-3.0	87	0.00
28 T	Bromochloromethane	0.334	0.366	-9.6	92	0.00
29 T	Tetrahydrofuran	0.078	0.085	-9.0	89	0.00
30 C	Chloroform	0.926	0.986	-6.5#	90	0.00
31 T	Cyclohexane	0.756	0.682	9.8	76	0.00
32 T	1,1,1-Trichloroethane	0.861	0.897	-4.2	84	0.00
33 S	1,2-Dichloroethane-d4	0.492	0.470	4.5	76	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.313	0.301	3.8	78	0.00
36 T	1,1-Dichloropropene	0.432	0.419	3.0	82	0.00
37 T	Ethyl Acetate	0.180	0.188	-4.4	86	0.00
38 T	Carbon Tetrachloride	0.521	0.513	1.5	81	0.00
39 T	Methylcyclohexane	0.554	0.520	6.1	77	0.00
40 TM	Benzene	1.254	1.270	-1.3	87	0.00
41 T	Methacrylonitrile	0.102	0.099	2.9	89	-0.04
42 TM	1,2-Dichloroethane	0.351	0.356	-1.4	87	0.00
43 T	Isopropyl Acetate	0.357	0.377	-5.6	89	0.00
44 TM	Trichloroethene	0.358	0.361	-0.8	87	0.00
45 C	1,2-Dichloropropane	0.273	0.281	-2.9#	89	0.00
46 T	Dibromomethane	0.177	0.187	-5.6	88	0.00
47 T	Bromodichloromethane	0.442	0.464	-5.0	89	0.00
48 T	Methyl methacrylate	0.162	0.166	-2.5	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071524\
 Data File : VY018810.D
 Acq On : 15 Jul 2024 11:08
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 15 13:12:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 09 14:28:51 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	90	0.00
50 S	Toluene-d8	1.186	1.100	7.3	74	0.00
51 T	4-Methyl-2-Pentanone	0.182	0.197	-8.2	89	0.00
52 CM	Toluene	0.833	0.836	-0.4#	86	0.00
53 T	t-1,3-Dichloropropene	0.382	0.417	-9.2	93	0.00
54 T	cis-1,3-Dichloropropene	0.448	0.473	-5.6	89	0.00
55 T	1,1,2-Trichloroethane	0.224	0.235	-4.9	90	0.00
56 T	Ethyl methacrylate	0.295	0.322	-9.2	92	0.00
57 T	1,3-Dichloropropane	0.371	0.395	-6.5	91	0.00
58 T	2-Chloroethyl Vinyl ether	0.141	0.166	-17.7	92	0.00
59 T	2-Hexanone	0.128	0.141	-10.2	90	0.00
60 T	Dibromochloromethane	0.308	0.337	-9.4	93	0.00
61 T	1,2-Dibromoethane	0.218	0.226	-3.7	86	0.00
62 S	4-Bromofluorobenzene	0.405	0.388	4.2	79	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
64 T	Tetrachloroethene	0.405	0.396	2.2	85	0.00
65 PM	Chlorobenzene	1.055	1.051	0.4	89	0.00
66 T	1,1,1,2-Tetrachloroethane	0.363	0.367	-1.1	90	0.00
67 C	Ethyl Benzene	1.848	1.775	4.0#	86	0.00
68 T	m/p-Xylenes	0.725	0.711	1.9	87	0.00
69 T	o-Xylene	0.682	0.668	2.1	86	0.00
70 T	Styrene	1.137	1.143	-0.5	89	0.00
71 P	Bromoform	0.198	0.216	-9.1	94	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
73 T	Isopropylbenzene	3.659	3.490	4.6	84	0.00
74 T	N-amyl acetate	0.748	0.765	-2.3	85	0.00
75 P	1,1,2,2-Tetrachloroethane	0.577	0.586	-1.6	90	0.00
76 T	1,2,3-Trichloropropane	0.392	0.474	-20.9	116	0.00
77 T	Bromobenzene	0.874	0.872	0.2	90	0.00
78 T	n-propylbenzene	4.346	4.182	3.8	85	0.00
79 T	2-Chlorotoluene	2.443	2.362	3.3	87	0.00
80 T	1,3,5-Trimethylbenzene	2.994	2.921	2.4	86	0.00
81 T	trans-1,4-Dichloro-2-butene	0.175	0.185	-5.7	89	0.00
82 T	4-Chlorotoluene	2.516	2.425	3.6	86	0.00
83 T	tert-Butylbenzene	2.764	2.694	2.5	87	0.00
84 T	1,2,4-Trimethylbenzene	2.975	2.861	3.8	86	0.00
85 T	sec-Butylbenzene	3.924	3.729	5.0	82	0.00
86 T	p-Isopropyltoluene	3.387	3.280	3.2	85	0.00
87 T	1,3-Dichlorobenzene	1.716	1.706	0.6	89	0.00
88 T	1,4-Dichlorobenzene	1.701	1.642	3.5	87	0.00
89 T	n-Butylbenzene	3.071	2.932	4.5	82	0.00
90 T	Hexachloroethane	0.630	0.636	-1.0	88	0.00
91 T	1,2-Dichlorobenzene	1.482	1.477	0.3	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.093	0.100	-7.5	88	0.00
93 T	1,2,4-Trichlorobenzene	0.905	0.976	-7.8	92	0.00
94 T	Hexachlorobutadiene	0.521	0.519	0.4	83	0.00
95 T	Naphthalene	1.552	1.707	-10.0	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071524\
Data File : VY018810.D
Acq On : 15 Jul 2024 11:08
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jul 15 13:12:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 09 14:28:51 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.774	0.830	-7.2	92	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6